


Data Path : Z:\VOASRV\HPCHEM1\MSVOA W\DATA\VW011419\
Data File : VW008237.D
Acq On : 14 Jan 2019 10:25
Operator : SY/VA
Sample : BFB
Misc : 5.00G/5ML/MSVOA W/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\82W010819S.M
Title : SW846 8260
Last Update : Wed Jan 09 04:40:48 2019



AutoFind: Scans 1757, 1758, 1759; Background Corrected with Scan 1748

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	19.7	21466	PASS
75	95	30	60	50.1	54536	PASS
95	95	100	100	100.0	108808	PASS
96	95	5	9	6.6	7182	PASS
173	174	0.00	2	0.0	0	PASS
174	95	50	100	78.3	85186	PASS
175	174	5	9	8.0	6786	PASS
176	174	95	101	97.0	82637	PASS
177	176	5	9	7.0	5774	PASS