

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW011422\
 Data File : VW021902.D
 Acq On : 14 Jan 2022 10:17
 Operator : SY/VA
 Sample : N1053-28
 Misc : 5.07g/10mL/MSVOA_W/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 VHBLK002

Quant Time: Jan 17 00:44:49 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM011122SMA.M
 Quant Title : SFAM01.0
 QLast Update : Mon Jan 17 00:44:09 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.842	114	126888	25.000	ug/L	# 0.00
28) Chlorobenzene-d5	11.628	117	127474	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.554	152	74640	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.349	65	58928	24.986	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	99.960%
7) Chloroethane-d5	2.879	69	45882	21.544	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	86.160%
11) 1,1-Dichloroethene-d2	4.013	63	67126	18.028	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	72.120%
21) 2-Butanone-d5	7.086	46	19141	57.501	ug/L	0.01
Spiked Amount	50.000	Range	20 - 135	Recovery	=	115.000%
24) Chloroform-d	7.647	84	99747	24.333	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	97.320%
26) 1,2-Dichloroethane-d4	8.305	65	52514	26.528	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	106.120%
32) Benzene-d6	8.275	84	202511	24.483	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	97.920%
36) 1,2-Dichloropropane-d6	9.274	67	55196	25.318	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	101.280%
41) Toluene-d8	10.323	98	195928	23.498	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	94.000%
43) trans-1,3-Dichloroprop...	10.579	79	22681	24.503	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	98.000%
47) 2-Hexanone-d5	10.927	63	18255	61.708	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	123.420%
56) 1,1,2,2-Tetrachloroeth...	12.688	84	45660	26.783	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	107.120%
66) 1,2-Dichlorobenzene-d4	13.853	152	76928	25.065	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	100.280%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW011422\
Data File : VW021902.D
Acq On : 14 Jan 2022 10:17
Operator : SY/VA
Sample : N1053-28
Misc : 5.07g/10mL/MSVOA_W/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
VHBLK002

Quant Time: Jan 17 00:44:49 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM011122SMA.M
Quant Title : SFAM01.0
QLast Update : Mon Jan 17 00:44:09 2022
Response via : Initial Calibration

