

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW011422\
 Data File : VW021904.D
 Acq On : 14 Jan 2022 11:05
 Operator : SY/VA
 Sample : N1084-18
 Misc : 5.00g/10mL/MSVOA_W/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 VHBLK002

Quant Time: Jan 17 00:45:07 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM011122SMA.M
 Quant Title : SFAM01.0
 QLast Update : Mon Jan 17 00:44:09 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.842	114	118278	25.000	ug/L	# 0.00
28) Chlorobenzene-d5	11.628	117	117552	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.554	152	66177	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.343	65	55868	25.413	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	101.640%
7) Chloroethane-d5	2.879	69	42891	21.606	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	86.440%
11) 1,1-Dichloroethene-d2	4.013	63	62992	18.150	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	72.600%
21) 2-Butanone-d5	7.086	46	19498	62.837	ug/L	0.01
Spiked Amount	50.000	Range	20 - 135	Recovery	=	125.680%
24) Chloroform-d	7.647	84	93844	24.559	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	98.240%
26) 1,2-Dichloroethane-d4	8.305	65	49414	26.780	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	107.120%
32) Benzene-d6	8.275	84	188969	24.774	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	99.080%
36) 1,2-Dichloropropane-d6	9.275	67	51039	25.387	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	101.560%
41) Toluene-d8	10.323	98	187551	24.391	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	97.560%
43) trans-1,3-Dichloroprop...	10.579	79	21210	24.848	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	99.400%
47) 2-Hexanone-d5	10.927	63	17710	64.919	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	129.840%
56) 1,1,2,2-Tetrachloroeth...	12.689	84	44238	28.140	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	112.560%
66) 1,2-Dichlorobenzene-d4	13.847	152	70895	26.054	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	104.200%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW011422\
Data File : VW021904.D
Acq On : 14 Jan 2022 11:05
Operator : SY/VA
Sample : N1084-18
Misc : 5.00g/10mL/MSVOA_W/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
VHBLK002

Quant Time: Jan 17 00:45:07 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM011122SMA.M
Quant Title : SFAM01.0
QLast Update : Mon Jan 17 00:44:09 2022
Response via : Initial Calibration

