

Data Path : Z:\VOASRV\HPCHEM1\MSVOA W\DATA\VW011519\
 Data File : VW008287.D
 Acq On : 15 Jan 2019 17:48
 Operator : SY/VA
 Sample : K1147-07
 Misc : 4.99G/5ML/MSVOA W/SOIL
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 TP-3

Quant Time: Jan 16 01:52:11 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\82W010819S.M
 Quant Title : SW846 8260
 QLast Update : Wed Jan 09 04:40:48 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.95	168	99312	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.84	114	174458	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.63	117	159617	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.56	152	75918	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.31	65	58675	52.57	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.14%	
35) Dibromofluoromethane	7.88	113	56027	55.06	ug/l	0.00
Spiked Amount 50.000			Recovery	=	110.12%	
50) Toluene-d8	10.32	98	215883	50.56	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.12%	
62) 4-Bromofluorobenzene	12.62	95	75084	49.05	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.10%	
Target Compounds						
16) Acetone	4.12	43	9740	29.928	ug/l	93
20) Methylene Chloride	4.92	84	10779	5.532	ug/l #	80

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA W\DATA\VW011519\
Data File : VW008287.D
Acq On : 15 Jan 2019 17:48
Operator : SY/VA
Sample : K1147-07
Misc : 4.99G/5ML/MSVOA W/SOIL
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
TP-3

Quant Time: Jan 16 01:52:11 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_W\METHOD\82W010819S.M
Quant Title : SW846 8260
QLast Update : Wed Jan 09 04:40:48 2019
Response via : Initial Calibration

