

Data Path : Z:\VOASRV\HPCHEM1\MSVOA W\DATA\VW011519\
 Data File : VW008297.D
 Acq On : 15 Jan 2019 22:16
 Operator : SY/VA
 Sample : K1012-09
 Misc : 6.74G/5ML/MSVOA W/SOIL
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 EO-01-011518-B

Quant Time: Jan 16 03:23:54 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\82W010819S.M
 Quant Title : SW846 8260
 QLast Update : Wed Jan 09 04:40:48 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.95	168	9693	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.84	114	15344	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.63	117	12888	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.56	152	4390	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.31	65	6322	58.03	ug/l	0.00
Spiked Amount 50.000			Recovery	=	116.06%	
35) Dibromofluoromethane	7.88	113	5412	60.47	ug/l	0.00
Spiked Amount 50.000			Recovery	=	120.94%	
50) Toluene-d8	10.32	98	17737	47.23	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.46%	
62) 4-Bromofluorobenzene	12.62	95	4908	36.46	ug/l	0.00
Spiked Amount 50.000			Recovery	=	72.92%	
Target Compounds						
					Qvalue	
20) Methylene Chloride	4.92	84	2803	23.594	ug/l #	80
52) Toluene	10.39	92	5125	18.719	ug/l	95
67) Ethyl Benzene	11.73	91	1529	3.061	ug/l #	80
68) m/p-Xylenes	11.84	106	2021	10.598	ug/l	91
69) o-Xylene	12.16	106	1004	5.686	ug/l	89
84) 1,2,4-Trimethylbenzene	13.26	105	1274	4.361	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA W\DATA\VW011519\
Data File : VW008297.D
Acq On : 15 Jan 2019 22:16
Operator : SY/VA
Sample : K1012-09
Misc : 6.74G/5ML/MSVOA W/SOIL
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
EO-01-011518-B

Quant Time: Jan 16 03:23:54 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_W\METHOD\82W010819S.M
Quant Title : SW846 8260
QLast Update : Wed Jan 09 04:40:48 2019
Response via : Initial Calibration

