

Data Path : Z:\VOASRV\HPCHEM1\MSVOA W\DATA\VW011520\
 Data File : VW014651.D
 Acq On : 15 Jan 2020 13:51
 Operator : SY/VA
 Sample : VIBLK33
 Misc : 5.00G/10ML/MSVOA W/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 VIBLK33

Quant Time: Jan 16 03:48:49 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\SOM2WLM122719S.M
 Quant Title : VOC Analysis
 QLast Update : Thu Jan 16 01:18:54 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	8.84	114	1112184	25.00	ug/L	0.00
28) Chlorobenzene-d5	11.63	117	997176	25.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	13.55	152	480176	25.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	2.35	65	342682	21.86	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	87.44%
7) Chloroethane-d5	2.89	69	268376	23.23	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	92.92%
10) 1,1-Dichloroethene-d2	4.01	63	490535	17.10	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	68.40%
20) 2-Butanone-d5	7.07	46	208587	55.91	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	111.82%
24) Chloroform-d	7.64	84	615469	24.05	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	96.20%
26) 1,2-Dichloroethane-d4	8.30	65	332683	25.05	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	100.20%
29) Benzene-d6	8.27	84	1384193	26.10	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	104.40%
33) 1,2-Dichloropropane-d6	9.27	67	418737	26.26	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	105.04%
37) Toluene-d8	10.32	98	1287310	26.78	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	107.12%
38) trans-1,3-Dichloropropene-	10.57	79	167426	25.42	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	101.68%
39) 2-Hexanone-d5	10.92	63	177943	65.27	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	130.54%
48) 1,1,2,2-Tetrachloroethane-	12.69	84	333577	27.86	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	111.44%
61) 1,2-Dichlorobenzene-d4	13.85	152	464333	29.35	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	117.40%

Target Compounds

					Ovalue
13) Acetone	4.12	43	18017	3.488 ug/L	68
34) Benzene	8.32	78	23408	0.390 ug/L	100
44) Toluene	10.38	91	142589	2.274 ug/L	99
54) m,p-Xylene	11.83	106	85551	3.279 ug/L	97
55) o-xylene	12.16	106	31861	1.293 ug/L	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA W\DATA\VW011520\
Data File : VW014651.D
Acq On : 15 Jan 2020 13:51
Operator : SY/VA
Sample : VIBLK33
Misc : 5.00G/10ML/MSVOA W/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
VIBLK33

Quant Time: Jan 16 03:48:49 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_W\METHOD\SOM2WLM122719S.M
Quant Title : VOC Analysis
QLast Update : Thu Jan 16 01:18:54 2020
Response via : Initial Calibration

