




Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW011522\
 Data File : VW021930.D
 Acq On : 15 Jan 2022 02:07
 Operator : SY/VA
 Sample : N1155-11
 Misc : 6.69g/10mL/MSVOA_W/SoIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 BGLW5

Manual IntegrationsAPPROVED

Quant Time: Jan 17 03:25:12 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM011122SMA.M
 Quant Title : SFAM01.0
 QLast Update : Mon Jan 17 03:23:20 2022
 Response via : Initial Calibration

Reviewed By :Semsettin Yesilyurt 01/17/2022
 Supervised By :Mahesh Dadoda 01/17/2022



TIC: VW021930.D\data.ms

(21) 2-Butanone-d5 (S)

7.153min (+ 0.073) 0.62 ug/L

response 259

Ion	Exp%	Act%
46.00	100.00	100.00
77.00	20.20	17.37
0.00	0.00	0.00
0.00	0.00	0.00




Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW011522\
 Data File : VW021930.D
 Acq On : 15 Jan 2022 02:07
 Operator : SY/VA
 Sample : N1155-11
 Misc : 6.69g/10mL/MSVOA_W/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 BGLW5

Manual IntegrationsAPPROVED

Quant Time: Jan 17 03:25:12 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM011122SMA.M
 Quant Title : SFAM01.0
 QLast Update : Mon Jan 17 03:23:20 2022
 Response via : Initial Calibration

Reviewed By :Semsettin Yesilyurt 01/17/2022
 Supervised By :Mahesh Dadoda 01/17/2022



TIC: VW021930.D\data.ms

(21) 2-Butanone-d5 (S)

7.080min (0.000) 51.64 ug/L m

response 21722

Ion	Exp%	Act%
46.00	100.00	100.00
77.00	20.20	0.21#
0.00	0.00	0.00
0.00	0.00	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW011522\
 Data File : VW021930.D
 Acq On : 15 Jan 2022 02:07
 Operator : SY/VA
 Sample : N1155-11
 Misc : 6.69g/10mL/MSVOA_W/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 BGLW5

Manual IntegrationsAPPROVED

Reviewed By :Semsettin Yesilyurt 01/17/2022
 Supervised By :Mahesh Dadoda 01/17/2022

Quant Time: Jan 17 03:25:12 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM011122SMA.M
 Quant Title : SFAM01.0
 QLast Update : Mon Jan 17 03:23:20 2022
 Response via : Initial Calibration

Compound	R.T.	Q	Ion	Response	Conc	Units	Dev(Min)

Internal Standards							
1) 1,4-Difluorobenzene	8.842	114		160347	25.000	ug/L	# 0.00
28) Chlorobenzene-d5	11.634	117		157739	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.554	152		86781	25.000	ug/L	0.00
System Monitoring Compounds							
4) Vinyl Chloride-d3	2.349	65		42400	14.227	ug/L	0.00
Spiked Amount 25.000	Range 30	- 150		Recovery =	56.920%		
7) Chloroethane-d5	2.879	69		32152	11.947	ug/L	0.00
Spiked Amount 25.000	Range 30	- 150		Recovery =	47.800%		
11) 1,1-Dichloroethene-d2	4.019	63		55509	11.797	ug/L	0.00
Spiked Amount 25.000	Range 45	- 110		Recovery =	47.200%		
21) 2-Butanone-d5	7.080	46		21722m	51.638	ug/L	0.00
Spiked Amount 50.000	Range 20	- 135		Recovery =	103.280%		
24) Chloroform-d	7.647	84		97262	18.776	ug/L	0.00
Spiked Amount 25.000	Range 40	- 150		Recovery =	75.120%		
26) 1,2-Dichloroethane-d4	8.305	65		51731	20.680	ug/L	0.00
Spiked Amount 25.000	Range 70	- 130		Recovery =	82.720%		
32) Benzene-d6	8.275	84		188242	18.391	ug/L	0.00
Spiked Amount 25.000	Range 20	- 135		Recovery =	73.560%		
36) 1,2-Dichloropropane-d6	9.275	67		53929	19.990	ug/L	0.00
Spiked Amount 25.000	Range 70	- 120		Recovery =	79.960%		
41) Toluene-d8	10.323	98		181769	17.617	ug/L	0.00
Spiked Amount 25.000	Range 30	- 130		Recovery =	70.480%		
43) trans-1,3-Dichloroprop...	10.579	79		20936	18.278	ug/L	0.00
Spiked Amount 25.000	Range 30	- 135		Recovery =	73.120%		
47) 2-Hexanone-d5	10.921	63		19030	51.986	ug/L	0.00
Spiked Amount 50.000	Range 20	- 135		Recovery =	103.980%		
56) 1,1,2,2-Tetrachloroeth...	12.689	84		47340	22.441	ug/L	0.00
Spiked Amount 25.000	Range 45	- 120		Recovery =	89.760%		
66) 1,2-Dichlorobenzene-d4	13.853	152		73021	20.464	ug/L	0.00
Spiked Amount 25.000	Range 75	- 120		Recovery =	81.840%		
Target Compounds							
13) Acetone	4.141	43		1954	5.622	ug/L	Qvalue 75

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW011522\
Data File : VW021930.D
Acq On : 15 Jan 2022 02:07
Operator : SY/VA
Sample : N1155-11
Misc : 6.69g/10mL/MSVOA_W/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
BGLW5

Manual IntegrationsAPPROVED

Reviewed By :Semsettin Yesilyurt 01/17/2022
Supervised By :Mahesh Dadoda 01/17/2022

Quant Time: Jan 17 03:25:12 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM011122SMA.M
Quant Title : SFAM01.0
QLast Update : Mon Jan 17 03:23:20 2022
Response via : Initial Calibration

