

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW011522\
 Data File : VW021933.D
 Acq On : 15 Jan 2022 03:19
 Operator : SY/VA
 Sample : N1155-04
 Misc : 6.51g/10mL/MSVOA_W/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 BGLT7

Quant Time: Jan 17 03:25:42 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM011122SMA.M
 Quant Title : SFAM01.0
 QLast Update : Mon Jan 17 03:23:20 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.842	114	170016	25.000	ug/L	# 0.00
28) Chlorobenzene-d5	11.628	117	164350	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.554	152	82337	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.355	65	44532	14.092	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	56.360%
7) Chloroethane-d5	2.892	69	36008	12.619	ug/L	0.01
Spiked Amount	25.000	Range	30 - 150	Recovery	=	50.480%
11) 1,1-Dichloroethene-d2	4.019	63	58497	11.725	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	46.920%
21) 2-Butanone-d5	7.086	46	22104	49.558	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	99.120%
24) Chloroform-d	7.653	84	101811	18.536	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	74.160%
26) 1,2-Dichloroethane-d4	8.305	65	56261	21.212	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	84.840%
32) Benzene-d6	8.275	84	198367	18.601	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	74.400%
36) 1,2-Dichloropropane-d6	9.275	67	56403	20.066	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	80.280%
41) Toluene-d8	10.323	98	189372	17.615	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	70.480%
43) trans-1,3-Dichloroprop...	10.579	79	22323	18.705	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	74.840%
47) 2-Hexanone-d5	10.921	63	17259	45.251	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	90.500%
56) 1,1,2,2-Tetrachloroeth...	12.689	84	51489	23.426	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	93.720%
66) 1,2-Dichlorobenzene-d4	13.853	152	64846	19.154	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	76.600%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW011522\
Data File : VW021933.D
Acq On : 15 Jan 2022 03:19
Operator : SY/VA
Sample : N1155-04
Misc : 6.51g/10mL/MSVOA_W/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
BGLT7

Quant Time: Jan 17 03:25:42 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM011122SMA.M
Quant Title : SFAM01.0
QLast Update : Mon Jan 17 03:23:20 2022
Response via : Initial Calibration

