

Data Path : Z:\VOASRV\HPCHEM1\MSVOA W\DATA\VW011619\
 Data File : VW008320.D
 Acq On : 16 Jan 2019 20:30
 Operator : SY/VA
 Sample : K1167-09
 Misc : 5.90G/5ML/MSVOA W/SOIL
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 CLOVER-YARD-A

Manual Integrations
 APPROVED

MMDadoda
 1/17/2019 11:33:39 AM

Quant Time: Jan 17 06:45:53 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\82W010819S.M
 Quant Title : SW846 8260
 QLast Update : Wed Jan 09 04:40:48 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.95	168	76108	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.84	114	140751	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.63	117	129788	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.56	152	61658	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.31	65	48924	57.20	ug/l	0.00
Spiked Amount 50.000			Recovery	=	114.40%	
35) Dibromofluoromethane	7.88	113	45886	55.90	ug/l	0.00
Spiked Amount 50.000			Recovery	=	111.80%	
50) Toluene-d8	10.32	98	182904	53.09	ug/l	0.00
Spiked Amount 50.000			Recovery	=	106.18%	
62) 4-Bromofluorobenzene	12.62	95	62329	50.47	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.94%	
Target Compounds						
16) Acetone	4.12	43	12239	63.852	ug/l	91
20) Methylene Chloride	4.90	84	10351	8.279	ug/l	93
25) 2-Butanone	7.17	43	3935	16.534	ug/l	96
67) Ethyl Benzene	11.73	91	7176	1.426	ug/l	93
70) Styrene	12.18	104	4636	1.664	ug/l	92
80) 1,3,5-Trimethylbenzene	12.90	105	7085m	1.781	ug/l	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA W\DATA\VW011619\
Data File : VW008320.D
Acq On : 16 Jan 2019 20:30
Operator : SY/VA
Sample : K1167-09
Misc : 5.90G/5ML/MSVOA W/SOIL
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
CLOVER-YARD-A

Manual Integrations
APPROVED

MMDadoda
1/17/2019 11:33:39 AM

Quant Time: Jan 17 06:45:53 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_W\METHOD\82W010819S.M
Quant Title : SW846 8260
QLast Update : Wed Jan 09 04:40:48 2019
Response via : Initial Calibration

