

Data Path : Z:\VOASRV\HPCHEM1\MSVOA W\DATA\VW011819\
 Data File : VW008353.D
 Acq On : 18 Jan 2019 13:55
 Operator : SY/VA
 Sample : K1188-03RE
 Misc : 6.83G/5ML/MSVOA W/SOIL
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 S3-FRONT-HOUSE-7RE

Quant Time: Jan 19 02:51:46 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\82W011719S.M
 Quant Title : SW846 8260
 QLast Update : Fri Jan 18 00:57:07 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.95	168	69784	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.84	114	131376	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.63	117	123546	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.56	152	53933	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.31	65	45813	63.11	ug/l	0.00
Spiked Amount 50.000			Recovery	=	126.22%	
35) Dibromofluoromethane	7.88	113	43505	51.88	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.76%	
50) Toluene-d8	10.32	98	174327	52.91	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.82%	
62) 4-Bromofluorobenzene	12.62	95	55283	46.79	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.58%	
Target Compounds						
16) Acetone	4.13	43	3546	21.034	ug/l #	87
20) Methylene Chloride	4.91	84	6865	5.627	ug/l #	93
52) Toluene	10.39	92	22640	9.555	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA W\DATA\VW011819\
Data File : VW008353.D
Acq On : 18 Jan 2019 13:55
Operator : SY/VA
Sample : K1188-03RE
Misc : 6.83G/5ML/MSVOA W/SOIL
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
S3-FRONT-HOUSE-7RE

Quant Time: Jan 19 02:51:46 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_W\METHOD\82W011719S.M
Quant Title : SW846 8260
QLast Update : Fri Jan 18 00:57:07 2019
Response via : Initial Calibration

