

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW012025\
 Data File : VW031581.D
 Acq On : 20 Jan 2025 14:37
 Operator : SY/MD
 Sample : Q1108-17
 Misc : 3.38g/10mL/MSVOA_W/SOIL/B
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 YE8E6

Quant Time: Jan 21 01:00:58 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM011025SMA.M
 Quant Title : SFAM01.0
 QLast Update : Sat Jan 18 00:35:19 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.837	114	1165744	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.629	117	889207	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.550	152	136184	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.363	65	394177	21.549	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	86.200%
7) Chloroethane-d5	2.899	69	273693	20.951	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	83.800%
11) 1,1-Dichloroethene-d2	4.015	65	175493	21.562	ug/L	-0.01
Spiked Amount	25.000	Range	45 - 110	Recovery	=	86.240%
21) 2-Butanone-d5	7.087	46	100925	34.322	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	68.640%
24) Chloroform-d	7.648	84	623807	20.677	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	82.720%
26) 1,2-Dichloroethane-d4	8.301	65	289929	19.525	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	78.080%
32) Benzene-d6	8.270	84	1359825	24.346	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	97.400%
36) 1,2-Dichloropropane-d6	9.270	67	385084	23.351	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	93.400%
41) Toluene-d8	10.319	98	1162019	23.046	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	92.200%
43) trans-1,3-Dichloroprop...	10.575	79	135050	19.942	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	79.760%
47) 2-Hexanone-d5	10.922	63	64842	38.529	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	77.060%
56) 1,1,2,2-Tetrachloroeth...	12.690	84	196943	19.528	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	78.120%
66) 1,2-Dichlorobenzene-d4	13.848	152	130891	27.403	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	109.600%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW012025\
Data File : VW031581.D
Acq On : 20 Jan 2025 14:37
Operator : SY/MD
Sample : Q1108-17
Misc : 3.38g/10mL/MSVOA_W/SOIL/B
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
YE8E6

Quant Time: Jan 21 01:00:58 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM011025SMA.M
Quant Title : SFAM01.0
QLast Update : Sat Jan 18 00:35:19 2025
Response via : Initial Calibration

