

Data Path : Z:\VOASRV\HPCHEM1\MSVOA W\DATA\VW012119\
 Data File : VW008379.D
 Acq On : 21 Jan 2019 17:00
 Operator : SY/VA
 Sample : K1009-07
 Misc : 6.34G/5ML/MSVOA W/SOIL
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 CL-01-011819-D

Quant Time: Jan 22 04:37:33 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\82W011719S.M
 Quant Title : SW846 8260
 QLast Update : Fri Jan 18 00:57:07 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.95	168	84209	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.84	114	159663	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.63	117	147585	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.56	152	53545	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.31	65	56027	63.96	ug/l	0.00
Spiked Amount	50.000		Recovery	=	127.92%	
35) Dibromofluoromethane	7.88	113	53376	52.37	ug/l	0.00
Spiked Amount	50.000		Recovery	=	104.74%	
50) Toluene-d8	10.33	98	206223	51.51	ug/l	0.00
Spiked Amount	50.000		Recovery	=	103.02%	
62) 4-Bromofluorobenzene	12.62	95	66077	46.01	ug/l	0.00
Spiked Amount	50.000		Recovery	=	92.02%	
Target Compounds						
16) Acetone	4.12	43	15304	75.229	ug/l	97
20) Methylene Chloride	4.90	84	5357	2.289	ug/l #	79
52) Toluene	10.39	92	4536208	1575.293	ug/l	96
67) Ethyl Benzene	11.73	91	306257	54.791	ug/l	97
68) m/p-Xylenes	11.84	106	638164	294.147	ug/l	98
69) o-Xylene	12.17	106	171408	84.566	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA W\DATA\VW012119\
Data File : VW008379.D
Acq On : 21 Jan 2019 17:00
Operator : SY/VA
Sample : K1009-07
Misc : 6.34G/5ML/MSVOA W/SOIL
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
CL-01-011819-D

Quant Time: Jan 22 04:37:33 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_W\METHOD\82W011719S.M
Quant Title : SW846 8260
QLast Update : Fri Jan 18 00:57:07 2019
Response via : Initial Calibration

