

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX012121\
 Data File : VX020703.D
 Acq On : 21 Jan 2021 17:59
 Operator : JC/SP
 Sample : VSTD0.0504
 Misc : 25.0mL/MSVOA_X/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 VSTD0.05203

Quant Time: Jan 22 01:26:18 2021
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\SFAMXSIM012121.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Fri Jan 22 01:23:42 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc Units	Dev(Min)
Internal Standards					
1) 1,4-Difluorobenzene	6.815	114	6733	0.50 ug/L	0.00
5) Chlorobenzene-d5	10.101	117	6090	0.50 ug/L	0.00
11) 1,4-Dichlorobenzene-d4	12.059	152	2827	0.50 ug/L	0.00
System Monitoring Compounds					
2) Vinyl Chloride-d3	1.387	65	335	0.05 ug/L	0.00
4) 1,2-Dichloroethane-d4	6.025	65	419	0.07 ug/L	0.01
7) 1,2-Dichloropropane-d6	7.361	67	301	0.05 ug/L	0.01
8) Toluene-d8	8.678	98	925	0.06 ug/L	0.00
10) 1,1,2,2-Tetrachloroeth...	11.219	84	255m	0.06 ug/L	0.00
Target Compounds					
3) Vinyl chloride	1.387	62	311	0.05 ug/L	Qvalue # 69
6) Trichloroethene	7.173	95	343	0.07 ug/L	94
9) 1,2-Dibromoethane	9.652	107	173	0.06 ug/L	Qvalue # 87
12) 1,2,3-Trichloropropane	11.271	75	147	0.05 ug/L	92
13) 1,2-Dibromo-3-chloropr...	12.967	75	27	0.05 ug/L	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX012121\
Data File : VX020703.D
Acq On : 21 Jan 2021 17:59
Operator : JC/SP
Sample : VSTD0.0504
Misc : 25.0mL/MSVOA_X/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
VSTD0.05203

Quant Time: Jan 22 01:26:18 2021
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\SFAMXSIM012121.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Fri Jan 22 01:23:42 2021
Response via : Initial Calibration

