

Data Path : Z:\VOASRV\HPCHEM1\MSVOA W\DATA\VW012320\
 Data File : VW014726.D
 Acq On : 23 Jan 2020 16:09
 Operator : SY/VA
 Sample : L1243-01
 Misc : 8.06G/5ML/MSVOA W/SOIL
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 CTY-SB-06-1.5-2.5

Quant Time: Jan 24 05:42:51 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\82W010320S.M
 Quant Title : SW846 8260
 QLast Update : Fri Jan 03 16:07:20 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.95	168	561338	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.84	114	943234	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.63	117	702069	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.55	152	209729	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.30	65	264841	53.96	ug/l	0.00
Spiked Amount 50.000			Recovery	=	107.92%	
35) Dibromofluoromethane	7.88	113	228631	42.41	ug/l	0.00
Spiked Amount 50.000			Recovery	=	84.82%	
50) Toluene-d8	10.32	98	871900	38.76	ug/l	0.00
Spiked Amount 50.000			Recovery	=	77.52%	
62) 4-Bromofluorobenzene	12.62	95	220315	28.87	ug/l	0.00
Spiked Amount 50.000			Recovery	=	57.74%	
Target Compounds						
16) Acetone	4.11	43	44482	32.098	ug/l	96
17) Carbon Disulfide	4.38	76	95858	5.693	ug/l	96
27) cis-1,2-Dichloroethene	7.16	96	6437	1.033	ug/l	92
44) Trichloroethene	9.09	130	18198	2.614	ug/l	89
64) Tetrachloroethene	10.86	164	26979	5.308	ug/l	92
80) 1,3,5-Trimethylbenzene	12.94	105	17104	1.324	ug/l	97
84) 1,2,4-Trimethylbenzene	13.25	105	29251	2.275	ug/l	97
95) Naphthalene	15.36	128	49378	7.019	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA W\DATA\VW012320\
Data File : VW014726.D
Acq On : 23 Jan 2020 16:09
Operator : SY/VA
Sample : L1243-01
Misc : 8.06G/5ML/MSVOA W/SOIL
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
CTY-SB-06-1.5-2.5

Quant Time: Jan 24 05:42:51 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_W\METHOD\82W010320S.M
Quant Title : SW846 8260
QLast Update : Fri Jan 03 16:07:20 2020
Response via : Initial Calibration

