

Data Path : Z:\VOASRV\HPCHEM1\MSVOA W\DATA\VW012320\
 Data File : VW014731.D
 Acq On : 23 Jan 2020 18:21
 Operator : SY/VA
 Sample : L1243-06
 Misc : 6.26G/5ML/MSVOA W/SOIL
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 CTY-SB-02-1-2

Quant Time: Jan 24 05:52:48 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\82W010320S.M
 Quant Title : SW846 8260
 QLast Update : Fri Jan 03 16:07:20 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.95	168	232928	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.84	114	346247	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.63	117	246202	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.55	152	52447	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.30	65	121182	59.50	ug/l	0.00
Spiked Amount	50.000		Recovery	=	119.00%	
35) Dibromofluoromethane	7.88	113	95520	48.27	ug/l	0.00
Spiked Amount	50.000		Recovery	=	96.54%	
50) Toluene-d8	10.32	98	361177	43.74	ug/l	0.00
Spiked Amount	50.000		Recovery	=	87.48%	
62) 4-Bromofluorobenzene	12.62	95	75215	26.85	ug/l	0.00
Spiked Amount	50.000		Recovery	=	53.70%	
Target Compounds						
16) Acetone	4.12	43	21083	36.663	ug/l	98
17) Carbon Disulfide	4.37	76	22308	3.193	ug/l	99
64) Tetrachloroethene	10.86	164	3527	1.979	ug/l	86
67) Ethyl Benzene	11.73	91	19267	2.120	ug/l	95
68) m/p-Xylenes	11.83	106	27474	7.897	ug/l	100
69) o-Xylene	12.16	106	8314	2.587	ug/l	85
84) 1,2,4-Trimethylbenzene	13.24	105	4480	1.393	ug/l	95
95) Naphthalene	15.36	128	2030	1.154	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA W\DATA\VW012320\
Data File : VW014731.D
Acq On : 23 Jan 2020 18:21
Operator : SY/VA
Sample : L1243-06
Misc : 6.26G/5ML/MSVOA W/SOIL
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
CTY-SB-02-1-2

Quant Time: Jan 24 05:52:48 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_W\METHOD\82W010320S.M
Quant Title : SW846 8260
QLast Update : Fri Jan 03 16:07:20 2020
Response via : Initial Calibration

