


Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW012419\
 Data File : VW008394.D
 Acq On : 24 Jan 2019 13:07
 Operator : SY/VA
 Sample : BFB
 Misc : 5.00G/5ML/MSVOA_W/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 BFB

Integration File: RTEINT.P

Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\82W012419S.M
 Title : SW846 8260
 Last Update : Fri Jan 18 00:57:07 2019



AutoFind: Scans 1757, 1758, 1759; Background Corrected with Scan 1749

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	19.9	23423	PASS
75	95	30	60	50.6	59733	PASS
95	95	100	100	100.0	117950	PASS
96	95	5	9	6.7	7958	PASS
173	174	0.00	2	0.5	433	PASS
174	95	50	100	78.4	92421	PASS
175	174	5	9	8.2	7586	PASS
176	174	95	101	97.2	89869	PASS
177	176	5	9	7.0	6263	PASS