

Data Path : Z:\VOASRV\HPCHEM1\MSVOA W\DATA\VW012519\
 Data File : VW008428.D
 Acq On : 25 Jan 2019 22:46
 Operator : SY/VA
 Sample : K1259-02
 Misc : 4.82G/5ML/MSVOA W/SOIL
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 FK-BPM-01-A01-B01-A01A-A01B

Quant Time: Jan 28 04:34:59 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\82W012419S.M
 Quant Title : SW846 8260
 QLast Update : Fri Jan 25 11:07:41 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.95	168	449559	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.84	114	719723	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.63	117	653605	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.56	152	330812	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.31	65	236684	53.59	ug/l	0.00
Spiked Amount 50.000			Recovery	=	107.18%	
35) Dibromofluoromethane	7.88	113	226564	53.84	ug/l	0.00
Spiked Amount 50.000			Recovery	=	107.68%	
50) Toluene-d8	10.32	98	863871	51.19	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.38%	
62) 4-Bromofluorobenzene	12.62	95	324080	49.96	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.92%	
Target Compounds						
16) Acetone	4.12	43	36577	36.880	ug/l	93
20) Methylene Chloride	4.91	84	11414	2.497	ug/l #	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA W\DATA\VW012519\
Data File : VW008428.D
Acq On : 25 Jan 2019 22:46
Operator : SY/VA
Sample : K1259-02
Misc : 4.82G/5ML/MSVOA W/SOIL
ALS Vial : 25 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
FK-BPM-01-A01-B01-A01A-A01B

Quant Time: Jan 28 04:34:59 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_W\METHOD\82W012419S.M
Quant Title : SW846 8260
QLast Update : Fri Jan 25 11:07:41 2019
Response via : Initial Calibration

