

Data Path : Z:\VOASRV\HPCHEM1\MSVOA W\DATA\VW013019\
 Data File : VW008461.D
 Acq On : 30 Jan 2019 13:11
 Operator : SY/VA
 Sample : K1272-01
 Misc : 5.02G/5ML/MSVOA W/SOIL
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 TS-01-012819

Quant Time: Jan 31 04:00:55 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\82W012419S.M
 Quant Title : SW846 8260
 QLast Update : Fri Jan 25 11:07:41 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.95	168	353531	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.84	114	562270	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.63	117	488544	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.56	152	212845	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.30	65	182616	52.58	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.16%	
35) Dibromofluoromethane	7.88	113	175979	53.53	ug/l	0.00
Spiked Amount 50.000			Recovery	=	107.06%	
50) Toluene-d8	10.32	98	651549	49.42	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.84%	
62) 4-Bromofluorobenzene	12.62	95	227519	44.90	ug/l	0.00
Spiked Amount 50.000			Recovery	=	89.80%	
Target Compounds						
16) Acetone	4.12	43	10140	13.001	ug/l	93
20) Methylene Chloride	4.90	84	12508	3.479	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA W\DATA\VW013019\
Data File : VW008461.D
Acq On : 30 Jan 2019 13:11
Operator : SY/VA
Sample : K1272-01
Misc : 5.02G/5ML/MSVOA W/SOIL
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
TS-01-012819

Quant Time: Jan 31 04:00:55 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_W\METHOD\82W012419S.M
Quant Title : SW846 8260
QLast Update : Fri Jan 25 11:07:41 2019
Response via : Initial Calibration

