

Data Path : Z:\voasrv\HPCHEM1\MSVOA W\Data\VW013019\
 Data File : VW008455.D
 Acq On : 30 Jan 2019 10:09
 Operator : SY/VA
 Sample : VSTDCCC050
 Misc : 5.00G/5ML/MSVOA W/SOIL
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_W
 LabSampleId :
 VSTDCCC050

Quant Time: Jan 31 01:37:53 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\82W012419S.M
 Quant Title : SW846 8260
 QLast Update : Fri Jan 25 11:07:41 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	77	0.00
2 T	Dichlorodifluoromethane	0.253	0.285	-12.6	74	0.00
3 P	Chloromethane	0.408	0.332	18.6	61	0.00
4 C	Vinyl Chloride	0.520	0.515	1.0#	70	0.00
5 T	Bromomethane	0.335	0.360	-7.5	80	0.00
6 T	Chloroethane	0.318	0.344	-8.2	78	0.00
7 T	Trichlorofluoromethane	0.389	0.417	-7.2	73	0.00
8 T	Diethyl Ether	0.249	0.230	7.6	68	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.452	0.498	-10.2	79	0.00
10 T	Methyl Iodide	0.714	0.775	-8.5	78	0.00
11 T	Tert butyl alcohol	0.040	0.034	15.0	66	0.00
12 CM	1,1-Dichloroethene	0.458	0.479	-4.6#	76	0.00
13 T	Acrolein	0.034	0.030	11.8	73	0.00
14 T	Allyl chloride	0.869	0.824	5.2	68	0.00
15 T	Acrylonitrile	0.108	0.092	14.8	62	0.00
16 T	Acetone	0.110	0.121	-10.0	83	0.00
17 T	Carbon Disulfide	1.462	1.411	3.5	69	0.00
18 T	Methyl Acetate	0.294	0.242	17.7	59	0.00
19 T	Methyl tert-butyl Ether	0.757	0.694	8.3	66	0.00
20 T	Methylene Chloride	0.508	0.473	6.9	74	0.00
21 T	trans-1,2-Dichloroethene	0.515	0.514	0.2	73	0.00
22 T	Diisopropyl ether	1.647	1.473	10.6	65	0.00
23 T	Vinyl Acetate	0.973	0.824	15.3	61	0.00
24 P	1,1-Dichloroethane	0.932	0.893	4.2	70	0.00
25 T	2-Butanone	0.153	0.141	7.8	66	0.00
26 T	2,2-Dichloropropane	0.630	0.647	-2.7	76	0.00
27 T	cis-1,2-Dichloroethene	0.561	0.555	1.1	72	0.00
28 T	Bromochloromethane	0.428	0.342	20.1#	65	0.00
29 T	Tetrahydrofuran	0.097	0.078	19.6	57	0.00
30 C	Chloroform	0.917	0.903	1.5#	71	0.00
31 T	Cyclohexane	0.964	0.854	11.4	67	0.00
32 T	1,1,1-Trichloroethane	0.792	0.811	-2.4	74	0.00
33 S	1,2-Dichloroethane-d4	0.491	0.442	10.0	69	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	74	0.00
35 S	Dibromofluoromethane	0.292	0.301	-3.1	74	0.00
36 T	1,1-Dichloropropene	0.472	0.492	-4.2	70	0.00
37 T	Ethyl Acetate	0.209	0.180	13.9	57	0.00
38 T	Carbon Tetrachloride	0.475	0.523	-10.1	75	0.00
39 T	Methylcyclohexane	0.620	0.620	0.0	68	0.00
40 TM	Benzene	1.311	1.325	-1.1	69	0.00
41 T	Methacrylonitrile	0.124	0.113	8.9	61	0.00
42 TM	1,2-Dichloroethane	0.380	0.364	4.2	67	0.00
43 T	Isopropyl Acetate	0.434	0.369	15.0	57	0.00
44 TM	Trichloroethene	0.356	0.385	-8.1	75	0.00
45 C	1,2-Dichloropropane	0.327	0.320	2.1#	67	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA W\Data\VW013019\
 Data File : VW008455.D
 Acq On : 30 Jan 2019 10:09
 Operator : SY/VA
 Sample : VSTDCCC050
 Misc : 5.00G/5ML/MSVOA W/SOIL
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_W
 LabSampleId :
 VSTDCCC050

Quant Time: Jan 31 01:37:53 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\82W012419S.M
 Quant Title : SW846 8260
 QLast Update : Fri Jan 25 11:07:41 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
46 T	Dibromomethane	0.168	0.161	4.2	66	0.00
47 T	Bromodichloromethane	0.445	0.445	0.0	69	0.00
48 T	Methyl methacrylate	0.212	0.176	17.0	57	0.00
49 T	1,4-Dioxane	0.003	0.003	0.0	72	0.00
50 S	Toluene-d8	1.172	1.199	-2.3	74	0.00
51 T	4-Methyl-2-Pentanone	0.214	0.182	15.0	58	0.00
52 CM	Toluene	0.865	0.862	0.3#	69	0.00
53 T	t-1,3-Dichloropropene	0.473	0.443	6.3	64	0.00
54 T	cis-1,3-Dichloropropene	0.537	0.524	2.4	67	0.00
55 T	1,1,2-Trichloroethane	0.233	0.221	5.2	66	0.00
56 T	Ethyl methacrylate	0.327	0.283	13.5	60	0.00
57 T	1,3-Dichloropropane	0.418	0.392	6.2	65	0.00
58 T	2-Chloroethyl Vinyl ether	0.159	0.136	14.5	67	0.00
59 T	2-Hexanone	0.151	0.144	4.6	65	0.00
60 T	Dibromochloromethane	0.300	0.292	2.7	67	0.00
61 T	1,2-Dibromoethane	0.231	0.217	6.1	65	0.00
62 S	4-Bromofluorobenzene	0.451	0.425	5.8	70	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	70	0.00
64 T	Tetrachloroethene	0.339	0.376	-10.9	73	0.00
65 PM	Chlorobenzene	1.021	1.093	-7.1	70	0.00
66 T	1,1,1,2-Tetrachloroethane	0.365	0.389	-6.6	69	0.00
67 C	Ethyl Benzene	1.876	1.996	-6.4#	69	0.00
68 T	m/p-Xylenes	0.728	0.769	-5.6	69	0.00
69 T	o-Xylene	0.692	0.727	-5.1	69	0.00
70 T	Styrene	1.142	1.174	-2.8	67	0.00
71 P	Bromoform	0.205	0.201	2.0	63	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	67	0.00
73 T	Isopropylbenzene	3.728	4.111	-10.3	69	0.00
74 T	N-amyl acetate	0.960	0.850	11.5	56	0.00
75 P	1,1,2,2-Tetrachloroethane	0.590	0.575	2.5	61	0.00
76 T	1,2,3-Trichloropropane	0.435	0.429	1.4	63	0.00
77 T	Bromobenzene	0.814	0.888	-9.1	69	0.00
78 T	n-propylbenzene	4.477	4.915	-9.8	68	0.00
79 T	2-Chlorotoluene	2.526	2.700	-6.9	67	0.00
80 T	1,3,5-Trimethylbenzene	3.116	3.379	-8.4	68	0.00
81 T	trans-1,4-Dichloro-2-butene	0.220	0.208	5.5	59	0.00
82 T	4-Chlorotoluene	2.690	2.816	-4.7	66	0.00
83 T	tert-Butylbenzene	2.773	3.087	-11.3	70	0.00
84 T	1,2,4-Trimethylbenzene	3.167	3.426	-8.2	68	0.00
85 T	sec-Butylbenzene	3.960	4.379	-10.6	69	0.00
86 T	p-Isopropyltoluene	3.468	3.892	-12.2	70	0.00
87 T	1,3-Dichlorobenzene	1.644	1.798	-9.4	68	0.00
88 T	1,4-Dichlorobenzene	1.634	1.765	-8.0	68	0.00
89 T	n-Butylbenzene	3.405	3.731	-9.6	68	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA W\Data\VW013019\
Data File : VW008455.D
Acq On : 30 Jan 2019 10:09
Operator : SY/VA
Sample : VSTDCCC050
Misc : 5.00G/5ML/MSVOA W/SOIL
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_W
LabSampleId :
VSTDCCC050

Quant Time: Jan 31 01:37:53 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\82W012419S.M
Quant Title : SW846 8260
QLast Update : Fri Jan 25 11:07:41 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
90 T	Hexachloroethane	0.678	0.757	-11.7	69	0.00
91 T	1,2-Dichlorobenzene	1.455	1.557	-7.0	67	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.112	0.100	10.7	57	0.00
93 T	1,2,4-Trichlorobenzene	1.038	1.125	-8.4	68	0.00
94 T	Hexachlorobutadiene	0.626	0.730	-16.6	75	0.00
95 T	Naphthalene	1.823	1.825	-0.1	62	0.00
96 T	1,2,3-Trichlorobenzene	0.881	0.931	-5.7	66	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6