

Data Path : Z:\VOASRV\HPCHEM1\MSVOA W\DATA\VW013020\
 Data File : VW014811.D
 Acq On : 30 Jan 2020 14:49
 Operator : SY/VA
 Sample : L1271-11
 Misc : 5.03G/10ML/MSVOA W/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 GAT02

Quant Time: Jan 31 01:43:21 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\SOM2WLM012720S.M
 Quant Title : VOC Analysis
 QLast Update : Fri Jan 31 01:08:17 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	8.84	114	945985	25.00	ug/L	0.00
28) Chlorobenzene-d5	11.63	117	818490	25.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	13.55	152	362669	25.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	2.35	65	399812	26.68	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	106.72%
7) Chloroethane-d5	2.89	69	343696	27.00	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	108.00%
10) 1,1-Dichloroethene-d2	4.01	63	589560	21.75	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	87.00%
20) 2-Butanone-d5	7.07	46	156397	36.76	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	73.52%
24) Chloroform-d	7.64	84	632951	24.24	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	96.96%
26) 1,2-Dichloroethane-d4	8.30	65	321979	22.90	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	91.60%
29) Benzene-d6	8.27	84	1308630	25.40	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	101.60%
33) 1,2-Dichloropropane-d6	9.27	67	349857	21.85	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	87.40%
37) Toluene-d8	10.32	98	1034208	22.54	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	90.16%
38) trans-1,3-Dichloropropene-	10.57	79	125876	19.73	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	78.92%
39) 2-Hexanone-d5	10.92	63	92141	30.66	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	61.32%
48) 1,1,2,2-Tetrachloroethane-	12.69	84	203792	16.22	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	64.88%
61) 1,2-Dichlorobenzene-d4	13.85	152	310558	21.85	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	87.40%

Target Compounds

					Ovalue
16) Methylene chloride	4.92	84	45942	2.874 ug/L	90
44) Toluene	10.38	91	97823	1.837 ug/L	97
56) Styrene	12.18	104	10587	0.306 ug/L	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA W\DATA\VW013020\
Data File : VW014811.D
Acq On : 30 Jan 2020 14:49
Operator : SY/VA
Sample : L1271-11
Misc : 5.03G/10ML/MSVOA W/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
GAT02

Quant Time: Jan 31 01:43:21 2020
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\SOM2WLM012720S.M
Quant Title : VOC Analysis
QLast Update : Fri Jan 31 01:08:17 2020
Response via : Initial Calibration

