

Data Path : Z:\VOASRV\HPCHEM1\MSVOA W\DATA\VW020119\
 Data File : VW008488.D
 Acq On : 01 Feb 2019 18:10
 Operator : SY/VA
 Sample : K1284-02
 Misc : 4.16G/5ML/MSVOA W/SOIL
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 FK-BPM-03-004-ABCD

Quant Time: Feb 02 02:22:42 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\82W012419S.M
 Quant Title : SW846 8260
 QLast Update : Fri Jan 25 11:07:41 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.95	168	363069	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.84	114	577622	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.63	117	532043	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.56	152	267282	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.30	65	191931	53.81	ug/l	0.00
Spiked Amount 50.000			Recovery	=	107.62%	
35) Dibromofluoromethane	7.88	113	183668	54.38	ug/l	0.00
Spiked Amount 50.000			Recovery	=	108.76%	
50) Toluene-d8	10.32	98	689591	50.91	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.82%	
62) 4-Bromofluorobenzene	12.62	95	261628	50.25	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.50%	
Target Compounds						
16) Acetone	4.12	43	17415	21.742	ug/l	98
20) Methylene Chloride	4.90	84	10302	2.791	ug/l	88

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA W\DATA\VW020119\
Data File : VW008488.D
Acq On : 01 Feb 2019 18:10
Operator : SY/VA
Sample : K1284-02
Misc : 4.16G/5ML/MSVOA W/SOIL
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
FK-BPM-03-004-ABCD

Quant Time: Feb 02 02:22:42 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_W\METHOD\82W012419S.M
Quant Title : SW846 8260
QLast Update : Fri Jan 25 11:07:41 2019
Response via : Initial Calibration

