

Data Path : Z:\VOASRV\HPCHEM1\MSVOA W\DATA\VW020119\
 Data File : VW008489.D
 Acq On : 01 Feb 2019 18:37
 Operator : SY/VA
 Sample : K1011-13
 Misc : 5.15G/5ML/MSVOA W/SOIL
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 OR-02-013119-A

Quant Time: Feb 02 03:21:52 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\82W012419S.M
 Quant Title : SW846 8260
 QLast Update : Fri Jan 25 11:07:41 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.95	168	354718	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.84	114	571220	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.63	117	472535	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.56	152	182023	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.30	65	184266	52.88	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.76%	
35) Dibromofluoromethane	7.88	113	179420	53.72	ug/l	0.00
Spiked Amount 50.000			Recovery	=	107.44%	
50) Toluene-d8	10.32	98	649684	48.51	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.02%	
62) 4-Bromofluorobenzene	12.62	95	209591	40.71	ug/l	0.00
Spiked Amount 50.000			Recovery	=	81.42%	
Target Compounds						
16) Acetone	4.12	43	54354	69.457	ug/l	98
20) Methylene Chloride	4.90	84	35345	9.799	ug/l	92
40) Benzene	8.32	78	32162	2.148	ug/l	98
52) Toluene	10.38	92	275110	27.835	ug/l	100
67) Ethyl Benzene	11.73	91	43393	2.447	ug/l	98
68) m/p-Xylenes	11.83	106	100345	14.586	ug/l	97
69) o-Xylene	12.16	106	37823	5.786	ug/l	95
84) 1,2,4-Trimethylbenzene	13.25	105	14578	1.265	ug/l	95
95) Naphthalene	15.37	128	45368	6.837	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA W\DATA\VW020119\
Data File : VW008489.D
Acq On : 01 Feb 2019 18:37
Operator : SY/VA
Sample : K1011-13
Misc : 5.15G/5ML/MSVOA W/SOIL
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
OR-02-013119-A

Quant Time: Feb 02 03:21:52 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_W\METHOD\82W012419S.M
Quant Title : SW846 8260
QLast Update : Fri Jan 25 11:07:41 2019
Response via : Initial Calibration

