

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW022022\
 Data File : VW022028.D
 Acq On : 02 Feb 2022 13:52
 Operator : SY/VA
 Sample : VIBLK
 Misc : 5.00g/10mL/MSVOA_W/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 VIBLK594

Quant Time: Feb 03 00:42:56 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM020122SMA.M
 Quant Title : SFAM01.0
 QLast Update : Thu Feb 03 00:41:48 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.841	114	154726	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.628	117	142636	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.554	152	71894	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.349	65	47769	25.631	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	102.520%
7) Chloroethane-d5	2.879	69	31430	21.905	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	87.640%
11) 1,1-Dichloroethene-d2	4.013	63	56908	17.282	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	69.120%
21) 2-Butanone-d5	7.086	46	17344	52.467	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	104.940%
24) Chloroform-d	7.653	84	84478	23.467	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	93.880%
26) 1,2-Dichloroethane-d4	8.305	65	44559	24.880	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	99.520%
32) Benzene-d6	8.274	84	169843	23.563	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	94.240%
36) 1,2-Dichloropropane-d6	9.274	67	46173	23.550	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	94.200%
41) Toluene-d8	10.323	98	166164	23.260	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	93.040%
43) trans-1,3-Dichloroprop...	10.579	79	18985	22.110	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	88.440%
47) 2-Hexanone-d5	10.920	63	15056	52.820	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	105.640%
56) 1,1,2,2-Tetrachloroeth...	12.688	84	38016	23.774	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	95.080%
66) 1,2-Dichlorobenzene-d4	13.853	152	60781	23.435	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	93.760%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW020222\
Data File : VW022028.D
Acq On : 02 Feb 2022 13:52
Operator : SY/VA
Sample : VIBLK
Misc : 5.00g/10mL/MSVOA_W/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
VIBLK594

Quant Time: Feb 03 00:42:56 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM020122SMA.M
Quant Title : SFAM01.0
QLast Update : Thu Feb 03 00:41:48 2022
Response via : Initial Calibration

