

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW022022\
 Data File : VW022033.D
 Acq On : 02 Feb 2022 16:01
 Operator : SY/VA
 Sample : VIBLK596
 Misc : 5.00g/10mL/MSVOA_W/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 VIBLK596

Quant Time: Feb 03 00:43:49 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM020122SMA.M
 Quant Title : SFAM01.0
 QLast Update : Thu Feb 03 00:41:48 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.842	114	161023	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.628	117	147376	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.554	152	75938	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.349	65	52682	27.161	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	108.640%
7) Chloroethane-d5	2.879	69	36825	24.662	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	98.640%
11) 1,1-Dichloroethene-d2	4.007	63	60893	17.769	ug/L	-0.01
Spiked Amount	25.000	Range	45 - 110	Recovery	=	71.080%
21) 2-Butanone-d5	7.080	46	18927	55.017	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	110.040%
24) Chloroform-d	7.647	84	88122	23.522	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	94.080%
26) 1,2-Dichloroethane-d4	8.305	65	46878	25.151	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	100.600%
32) Benzene-d6	8.275	84	179027	24.038	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	96.160%
36) 1,2-Dichloropropane-d6	9.274	67	47919	23.655	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	94.600%
41) Toluene-d8	10.323	98	177063	23.989	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	95.960%
43) trans-1,3-Dichloroprop...	10.573	79	20003	22.546	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	90.200%
47) 2-Hexanone-d5	10.921	63	15768	53.538	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	107.080%
56) 1,1,2,2-Tetrachloroeth...	12.688	84	39950	24.180	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	96.720%
66) 1,2-Dichlorobenzene-d4	13.853	152	64424	23.517	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	94.080%
Target Compounds					Qvalue	
53) m,p-Xylene	11.835	106	4473	1.116	ug/L	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW020222\
Data File : VW022033.D
Acq On : 02 Feb 2022 16:01
Operator : SY/VA
Sample : VIBLK596
Misc : 5.00g/10mL/MSVOA_W/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
VIBLK596

Quant Time: Feb 03 00:43:49 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM020122SMA.M
Quant Title : SFAM01.0
QLast Update : Thu Feb 03 00:41:48 2022
Response via : Initial Calibration

