

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW022022\
 Data File : VW022038.D
 Acq On : 02 Feb 2022 18:02
 Operator : SY/VA
 Sample : N1363-08
 Misc : 6.44g/10mL/MSVOA_W/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 BGLZ4

Quant Time: Feb 03 00:44:38 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM020122SMA.M
 Quant Title : SFAM01.0
 QLast Update : Thu Feb 03 00:41:48 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.842	114	170947	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.628	117	156548	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.554	152	71605	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.349	65	50116	24.338	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	97.360%
7) Chloroethane-d5	2.885	69	36855	23.249	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	93.000%
11) 1,1-Dichloroethene-d2	4.013	63	58629	16.116	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	64.480%
21) 2-Butanone-d5	7.080	46	17823	48.800	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	97.600%
24) Chloroform-d	7.647	84	84841	21.332	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	85.320%
26) 1,2-Dichloroethane-d4	8.305	65	45762	23.127	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	92.520%
32) Benzene-d6	8.275	84	172135	21.759	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	87.040%
36) 1,2-Dichloropropane-d6	9.268	67	46976	21.831	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	87.320%
41) Toluene-d8	10.323	98	168957	21.549	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	86.200%
43) trans-1,3-Dichloroprop...	10.573	79	19678	20.880	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	83.520%
47) 2-Hexanone-d5	10.921	63	16273	52.016	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	104.040%
56) 1,1,2,2-Tetrachloroeth...	12.688	84	37788	21.531	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	86.120%
66) 1,2-Dichlorobenzene-d4	13.853	152	54232	20.995	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	83.960%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW020222\
Data File : VW022038.D
Acq On : 02 Feb 2022 18:02
Operator : SY/VA
Sample : N1363-08
Misc : 6.44g/10mL/MSVOA_W/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
BGLZ4

Quant Time: Feb 03 00:44:38 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM020122SMA.M
Quant Title : SFAM01.0
QLast Update : Thu Feb 03 00:41:48 2022
Response via : Initial Calibration

