

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW022022\
 Data File : VW022041.D
 Acq On : 02 Feb 2022 19:14
 Operator : SY/VA
 Sample : N1363-11
 Misc : 4.70g/10mL/MSVOA_W/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 BGLZ7

Quant Time: Feb 03 00:45:12 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM020122SMA.M
 Quant Title : SFAM01.0
 QLast Update : Thu Feb 03 00:41:48 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.842	114	184963	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.628	117	171574	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.554	152	86016	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.349	65	49456	22.198	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	88.800%
7) Chloroethane-d5	2.879	69	36891	21.508	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	86.040%
11) 1,1-Dichloroethene-d2	4.013	63	60242	15.304	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	61.200%
21) 2-Butanone-d5	7.074	46	18331	46.387	ug/L	-0.01
Spiked Amount	50.000	Range	20 - 135	Recovery	=	92.780%
24) Chloroform-d	7.653	84	89656	20.834	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	83.320%
26) 1,2-Dichloroethane-d4	8.305	65	47714	22.286	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	89.160%
32) Benzene-d6	8.275	84	180471	20.815	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	83.240%
36) 1,2-Dichloropropane-d6	9.274	67	49270	20.892	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	83.560%
41) Toluene-d8	10.323	98	173940	20.242	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	80.960%
43) trans-1,3-Dichloroprop...	10.579	79	20508	19.855	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	79.400%
47) 2-Hexanone-d5	10.920	63	16388	47.796	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	95.600%
56) 1,1,2,2-Tetrachloroeth...	12.688	84	40621	21.118	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	84.480%
66) 1,2-Dichlorobenzene-d4	13.847	152	62745	20.221	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	80.880%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW020222\
Data File : VW022041.D
Acq On : 02 Feb 2022 19:14
Operator : SY/VA
Sample : N1363-11
Misc : 4.70g/10mL/MSVOA_W/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
BGLZ7

Quant Time: Feb 03 00:45:12 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM020122SMA.M
Quant Title : SFAM01.0
QLast Update : Thu Feb 03 00:41:48 2022
Response via : Initial Calibration

