

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW022022\
 Data File : VW022052.D
 Acq On : 03 Feb 2022 01:01
 Operator : SY/VA
 Sample : N1364-03
 Misc : 5.69g/10mL/MSVOA_W/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 BGM08

Quant Time: Feb 03 02:21:56 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM020122SMA.M
 Quant Title : SFAM01.0
 QLast Update : Thu Feb 03 01:57:48 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.842	114	115914	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.628	117	111948	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.554	152	51976	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.349	65	32198	23.061	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	92.240%
7) Chloroethane-d5	2.885	69	24790	23.063	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	92.240%
11) 1,1-Dichloroethene-d2	4.019	63	39620	16.061	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	64.240%
21) 2-Butanone-d5	7.080	46	12230	49.384	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	98.760%
24) Chloroform-d	7.653	84	61084	22.650	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	90.600%
26) 1,2-Dichloroethane-d4	8.305	65	33950	25.303	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	101.200%
32) Benzene-d6	8.275	84	121810	21.532	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	86.120%
36) 1,2-Dichloropropane-d6	9.274	67	34340	22.316	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	89.280%
41) Toluene-d8	10.323	98	118213	21.084	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	84.320%
43) trans-1,3-Dichloroprop...	10.579	79	13667	20.279	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	81.120%
47) 2-Hexanone-d5	10.920	63	10891	48.682	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	97.360%
56) 1,1,2,2-Tetrachloroeth...	12.688	84	28157	22.435	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	89.760%
66) 1,2-Dichlorobenzene-d4	13.853	152	41989	22.394	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	89.560%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW020222\
Data File : VW022052.D
Acq On : 03 Feb 2022 01:01
Operator : SY/VA
Sample : N1364-03
Misc : 5.69g/10mL/MSVOA_W/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
BGM08

Quant Time: Feb 03 02:21:56 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM020122SMA.M
Quant Title : SFAM01.0
QLast Update : Thu Feb 03 01:57:48 2022
Response via : Initial Calibration

