

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW020222\
 Data File : VW022039.D
 Acq On : 02 Feb 2022 18:26
 Operator : SY/VA
 Sample : N1363-09
 Misc : 6.13g/10mL/MSVOA_W/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 BGLZ5

Quant Time: Feb 03 00:44:48 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM020122SMA.M
 Quant Title : SFAM01.0
 QLast Update : Thu Feb 03 00:41:48 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.848	114	188687	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.628	117	172993	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.554	152	84120	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.349	65	49475	21.768	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	87.080%
7) Chloroethane-d5	2.885	69	36459	20.837	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	83.360%
11) 1,1-Dichloroethene-d2	4.013	63	59193	14.741	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	58.960%
21) 2-Butanone-d5	7.080	46	16901	41.925	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	83.840%
24) Chloroform-d	7.647	84	87055	19.831	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	79.320%
26) 1,2-Dichloroethane-d4	8.305	65	45982	21.053	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	84.200%
32) Benzene-d6	8.275	84	177202	20.270	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	81.080%
36) 1,2-Dichloropropane-d6	9.274	67	47641	20.035	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	80.160%
41) Toluene-d8	10.323	98	172319	19.889	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	79.560%
43) trans-1,3-Dichloroprop...	10.579	79	19644	18.863	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	75.440%
47) 2-Hexanone-d5	10.921	63	15696	45.402	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	90.800%
56) 1,1,2,2-Tetrachloroeth...	12.689	84	39229	20.227	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	80.920%
66) 1,2-Dichlorobenzene-d4	13.847	152	62845	20.709	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	82.840%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW020222\
Data File : VW022039.D
Acq On : 02 Feb 2022 18:26
Operator : SY/VA
Sample : N1363-09
Misc : 6.13g/10mL/MSVOA_W/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
BGLZ5

Quant Time: Feb 03 00:44:48 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM020122SMA.M
Quant Title : SFAM01.0
QLast Update : Thu Feb 03 00:41:48 2022
Response via : Initial Calibration

