

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW0220222\
 Data File : VW022051.D
 Acq On : 03 Feb 2022 00:37
 Operator : SY/VA
 Sample : N1364-02
 Misc : 5.57g/10mL/MSVOA_W/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 BGM07

Quant Time: Feb 03 02:21:46 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM020122SMA.M
 Quant Title : SFAM01.0
 QLast Update : Thu Feb 03 01:57:48 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.842	114	164359	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.628	117	156190	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.560	152	75796	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.349	65	48134	24.313	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	97.240%
7) Chloroethane-d5	2.879	69	32438	21.283	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	85.120%
11) 1,1-Dichloroethene-d2	4.013	63	56587	16.178	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	64.720%
21) 2-Butanone-d5	7.074	46	21672	61.717	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	123.440%
24) Chloroform-d	7.653	84	90659	23.708	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	94.840%
26) 1,2-Dichloroethane-d4	8.305	65	50665	26.631	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	106.520%
32) Benzene-d6	8.275	84	181644	23.013	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	92.040%
36) 1,2-Dichloropropane-d6	9.274	67	50898	23.708	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	94.840%
41) Toluene-d8	10.323	98	176103	22.512	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	90.040%
43) trans-1,3-Dichloroprop...	10.579	79	21574	22.944	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	91.760%
47) 2-Hexanone-d5	10.920	63	19233	61.618	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	123.240%
56) 1,1,2,2-Tetrachloroeth...	12.688	84	46124	26.341	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	105.360%
66) 1,2-Dichlorobenzene-d4	13.853	152	66472	24.310	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	97.240%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW020222\
Data File : VW022051.D
Acq On : 03 Feb 2022 00:37
Operator : SY/VA
Sample : N1364-02
Misc : 5.57g/10mL/MSVOA_W/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
BGM07

Quant Time: Feb 03 02:21:46 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM020122SMA.M
Quant Title : SFAM01.0
QLast Update : Thu Feb 03 01:57:48 2022
Response via : Initial Calibration

