

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW020322\
 Data File : VW022073.D
 Acq On : 03 Feb 2022 17:15
 Operator : SY/VA
 Sample : N1363-20
 Misc : 5.00g/10mL/MSVOA_W/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 VHBLK001

Quant Time: Feb 04 00:59:31 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM020122SMA.M
 Quant Title : SFAM01.0
 QLast Update : Thu Feb 03 01:57:48 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.842	114	143729	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.628	117	132269	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.554	152	63891	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.343	65	51231	29.591	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	118.360%
7) Chloroethane-d5	2.879	69	39133	29.361	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	117.440%
11) 1,1-Dichloroethene-d2	4.013	63	52976	17.319	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	69.280%
21) 2-Butanone-d5	7.074	46	16378	53.335	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	106.680%
24) Chloroform-d	7.647	84	81294	24.311	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	97.240%
26) 1,2-Dichloroethane-d4	8.305	65	43498	26.145	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	104.600%
32) Benzene-d6	8.275	84	161572	24.173	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	96.680%
36) 1,2-Dichloropropane-d6	9.268	67	44586	24.523	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	98.080%
41) Toluene-d8	10.323	98	157951	23.843	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	95.360%
43) trans-1,3-Dichloroprop...	10.579	79	17845	22.411	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	89.640%
47) 2-Hexanone-d5	10.927	63	14373	54.376	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	108.760%
56) 1,1,2,2-Tetrachloroeth...	12.688	84	36209	24.419	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	97.680%
66) 1,2-Dichlorobenzene-d4	13.853	152	58039	25.181	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	100.720%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW020322\
Data File : VW022073.D
Acq On : 03 Feb 2022 17:15
Operator : SY/VA
Sample : N1363-20
Misc : 5.00g/10mL/MSVOA_W/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
VHBLK001

Quant Time: Feb 04 00:59:31 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM020122SMA.M
Quant Title : SFAM01.0
QLast Update : Thu Feb 03 01:57:48 2022
Response via : Initial Calibration

