

Data Path : Z:\VOASRV\HPCHEM1\MSVOA W\DATA\VW020419\
 Data File : VW008514.D
 Acq On : 04 Feb 2019 20:50
 Operator : SY/VA
 Sample : K1346-03
 Misc : 6.15G/5ML/MSVOA W/SOIL
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 CL-02-020419-B

Manual Integrations
 APPROVED

MMDadoda
 2/5/2019 10:43:24 AM

Quant Time: Feb 05 04:16:02 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\82W012419S.M
 Quant Title : SW846 8260
 QLast Update : Fri Jan 25 11:07:41 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.95	168	342933	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.84	114	554055	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.63	117	461149	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.56	152	169457	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.30	65	188608	55.98	ug/l	0.00
Spiked Amount 50.000			Recovery	=	111.96%	
35) Dibromofluoromethane	7.88	113	175976	54.32	ug/l	0.00
Spiked Amount 50.000			Recovery	=	108.64%	
50) Toluene-d8	10.32	98	634848	48.87	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.74%	
62) 4-Bromofluorobenzene	12.62	95	200475	40.15	ug/l	0.00
Spiked Amount 50.000			Recovery	=	80.30%	
Target Compounds						
16) Acetone	4.12	43	66297	87.630	ug/l	99
20) Methylene Chloride	4.92	84	10275m	2.947	ug/l	
84) 1,2,4-Trimethylbenzene	13.26	105	26204	2.441	ug/l	96
95) Naphthalene	15.37	128	24167	3.912	ug/l #	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA W\DATA\VW020419\
Data File : VW008514.D
Acq On : 04 Feb 2019 20:50
Operator : SY/VA
Sample : K1346-03
Misc : 6.15G/5ML/MSVOA W/SOIL
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
CL-02-020419-B

Manual Integrations
APPROVED

MMDadoda
2/5/2019 10:43:24 AM

Quant Time: Feb 05 04:16:02 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_W\METHOD\82W012419S.M
Quant Title : SW846 8260
QLast Update : Fri Jan 25 11:07:41 2019
Response via : Initial Calibration

