

Data File : VW014883.D
Acq On : 05 Feb 2020 22:25
Operator : SY/VA
Sample : L1324-13
Misc : 5.03G/10ML/MSVOA_W/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :

Quant Time: Feb 06 03:55:52 2020
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\SOM2WLM020520S.M
Quant Title : VOC Analysis
QLast Update : Thu Feb 06 02:59:09 2020
Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	8.84	114	934966	25.00	ug/L	0.00
28) Chlorobenzene-d5	11.63	117	822230	25.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	13.55	152	285908	25.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	2.36	65	339189	21.91	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	87.64%
7) Chloroethane-d5	2.89	69	332836	26.86	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	107.44%
10) 1,1-Dichloroethene-d2	4.02	63	490398	17.34	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	69.36%
20) 2-Butanone-d5	7.07	46	231234	53.23	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	106.46%
24) Chloroform-d	7.65	84	747690	28.17	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	112.68%
26) 1,2-Dichloroethane-d4	8.31	65	434824m	29.60	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	118.40%
29) Benzene-d6	8.27	84	1461455	27.64	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	110.56%
33) 1,2-Dichloropropane-d6	9.27	67	517130	30.01	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	120.04%#
37) Toluene-d8	10.32	98	1219143	26.20	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	104.80%
38) trans-1,3-Dichloropropene-	10.57	79	164744	24.21	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	96.84%
39) 2-Hexanone-d5	10.92	63	173360	55.20	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	110.40%
48) 1,1,2,2-Tetrachloroethane-	12.69	84	362987	27.97	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	111.88%
61) 1,2-Dichlorobenzene-d4	13.85	152	292493	26.38	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	105.52%

Target Compounds

					Qvalue
13) Acetone	4.12	43	20328	4.534 ug/L	97
16) Methylene chloride	4.92	84	41751	2.451 ug/L	98
44) Toluene	10.38	91	83908	1.462 ug/L	99
53) Ethylbenzene	11.73	91	145908	2.324 ug/L	98
54) m,p-Xylene	11.84	106	20766	0.905 ug/L	99
55) o-xylene	12.16	106	27135	1.243 ug/L	100
56) Styrene	12.18	104	54188	1.469 ug/L	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data File : VW014883.D
Acq On : 05 Feb 2020 22:25
Operator : SY/VA
Sample : L1324-13
Misc : 5.03G/10ML/MSVOA_W/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :

Quant Time: Feb 06 03:55:52 2020
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\SOM2WLM020520S.M
Quant Title : VOC Analysis
QLast Update : Thu Feb 06 02:59:09 2020
Response via : Initial Calibration

