

Data Path : Z:\VOASRV\HPCHEM1\MSVOA W\DATA\VW020619\
 Data File : VW008540.D
 Acq On : 06 Feb 2019 20:48
 Operator : SY/VA
 Sample : K1389-15
 Misc : 5.10G/5ML/MSVOA W/SOIL
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 M-5-D

Quant Time: Feb 07 03:52:55 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\82W012419S.M
 Quant Title : SW846 8260
 QLast Update : Fri Jan 25 11:07:41 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.95	168	512952	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.84	114	819836	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.63	117	765551	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.56	152	392963	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.30	65	276011	54.77	ug/l	0.00
Spiked Amount 50.000			Recovery	=	109.54%	
35) Dibromofluoromethane	7.88	113	261774	54.61	ug/l	0.00
Spiked Amount 50.000			Recovery	=	109.22%	
50) Toluene-d8	10.32	98	980266	50.99	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.98%	
62) 4-Bromofluorobenzene	12.62	95	374122	50.63	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.26%	
Target Compounds						
12) 1,1-Dichloroethene	4.04	96	14900	3.173	ug/l #	77
16) Acetone	4.12	43	64987	57.427	ug/l	99
52) Toluene	10.38	92	22951	1.618	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA W\DATA\VW020619\
Data File : VW008540.D
Acq On : 06 Feb 2019 20:48
Operator : SY/VA
Sample : K1389-15
Misc : 5.10G/5ML/MSVOA W/SOIL
ALS Vial : 25 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
M-5-D

Quant Time: Feb 07 03:52:55 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_W\METHOD\82W012419S.M
Quant Title : SW846 8260
QLast Update : Fri Jan 25 11:07:41 2019
Response via : Initial Calibration

