

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW020624\
 Data File : VW027967.D
 Acq On : 06 Feb 2024 10:57
 Operator : SY/MD
 Sample : P1286-04
 Misc : 5.09g/10mL/MSVOA_W/SOIL/A
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 VHBLK001

Quant Time: Feb 06 23:52:05 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM020524SMA.M
 Quant Title : SFAM01.0
 QLast Update : Tue Feb 06 23:51:08 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.843	114	324973	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.629	117	297321	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.556	152	141526	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.356	65	90981	23.904	ug/L	-0.01
Spiked Amount	25.000	Range	30 - 150	Recovery	=	95.600%
7) Chloroethane-d5	2.899	69	74320	25.744	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	102.960%
11) 1,1-Dichloroethene-d2	4.021	65	39760	24.414	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	97.640%
21) 2-Butanone-d5	7.099	46	26147	44.212	ug/L	0.01
Spiked Amount	50.000	Range	20 - 135	Recovery	=	88.420%
24) Chloroform-d	7.648	84	180545	23.760	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	95.040%
26) 1,2-Dichloroethane-d4	8.306	65	87686	25.551	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	102.200%
32) Benzene-d6	8.276	84	389391	24.478	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	97.920%
36) 1,2-Dichloropropane-d6	9.276	67	102392	23.740	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	94.960%
41) Toluene-d8	10.324	98	365747	24.030	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	96.120%
43) trans-1,3-Dichloroprop...	10.574	79	41890	23.617	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	94.480%
47) 2-Hexanone-d5	10.922	63	29118	49.633	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	99.260%
56) 1,1,2,2-Tetrachloroeth...	12.690	84	75585	23.447	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	93.800%
66) 1,2-Dichlorobenzene-d4	13.854	152	124142	25.807	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	103.240%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW020624\
Data File : VW027967.D
Acq On : 06 Feb 2024 10:57
Operator : SY/MD
Sample : P1286-04
Misc : 5.09g/10mL/MSVOA_W/SOIL/A
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
VHBLK001

Quant Time: Feb 06 23:52:05 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM020524SMA.M
Quant Title : SFAM01.0
QLast Update : Tue Feb 06 23:51:08 2024
Response via : Initial Calibration

