


Data Path : Z:\V0ASRV\HPCHEM1\MSVOA W\DATA\VW020819\
Data File : VW008568.D
Acq On : 08 Feb 2019 10:03
Operator : SY/VA
Sample : BFB
Misc : 5.00G/5ML/MSVOA W/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\V0ASRV\HPCHEM1\MSVOA_W\METHOD\82W012419S.M
Title : SW846 8260
Last Update : Fri Jan 25 11:07:41 2019



AutoFind: Scans 1757, 1758, 1759; Background Corrected with Scan 1748

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	19.1	30037	PASS
75	95	30	60	50.6	79536	PASS
95	95	100	100	100.0	157269	PASS
96	95	5	9	6.4	10053	PASS
173	174	0.00	2	0.0	0	PASS
174	95	50	100	77.9	122458	PASS
175	174	5	9	8.2	10008	PASS
176	174	95	101	95.5	117008	PASS
177	176	5	9	6.6	7741	PASS