



Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW020821\
 Data File : VW018043.D
 Acq On : 08 Feb 2021 09:00
 Operator : SY/VA
 Sample : BFB37
 Misc : 5.00G/10ML/MSVOA_W/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 BFB37

Integration File: LSCINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SOM2WLM012921S.M
 Title : VOC Analysis
 Last Update : Tue Feb 09 03:23:41 2021



AutoFind: Scans 1756, 1757, 1758; Background Corrected with Scan 1749

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	16.2	10649	PASS
75	95	30	80	48.3	31691	PASS
95	95	100	100	100.0	65661	PASS
96	95	5	9	6.5	4288	PASS
173	174	0.00	2	0.0	0	PASS
174	95	50	120	81.2	53320	PASS
175	174	5	9	7.4	3947	PASS
176	174	95	101	96.0	51181	PASS
177	176	5	9	7.6	3867	PASS

m/z	Abundance
36.00	637.0
37.00	2646.0
38.00	2239.0
39.00	758.0
40.00	187.0
41.10	52.0
42.90	104.0
43.90	506.0
45.10	619.0
47.00	662.0
48.10	303.0
49.00	2275.0
50.00	9840.0
51.00	3004.0
52.20	84.0
55.00	226.0
56.00	847.0
57.00	1404.0
59.00	307.0
60.00	532.0
61.00	2640.0
62.00	2549.0
63.00	1860.0
64.00	201.0
65.00	183.0
66.80	117.0
68.00	5312.0
69.00	5704.0
70.00	408.0
72.10	293.0
73.00	3954.0
74.00	9926.0
75.00	28088.0
76.00	2261.0
76.90	316.0
78.00	230.0
78.90	1500.0
80.00	510.0
81.00	1709.0
82.00	416.0
83.00	98.0
84.90	60.0
86.90	2296.0
88.00	2151.0
89.00	114.0
90.90	191.0
92.00	1722.0
93.00	2391.0
94.00	6546.0
95.00	57872.0
96.00	3958.0
97.00	188.0
102.80	305.0
103.10	298.0
103.90	422.0
104.90	208.0
106.00	330.0
106.90	69.0
110.00	170.0
110.90	185.0
111.80	144.0
112.80	59.0
114.90	250.0
115.90	241.0
117.00	485.0
118.00	524.0
119.00	593.0
119.70	68.0
121.00	75.0
125.00	552.0
125.70	237.0
127.90	209.0
128.90	115.0
129.90	215.0
131.00	168.0
133.10	2096.0
134.00	351.0
135.00	260.0
136.90	123.0
140.90	545.0
142.00	77.0
143.00	674.0
145.10	85.0
147.00	315.0
148.00	74.0
149.00	226.0
154.90	188.0
156.90	108.0
159.00	103.0
160.80	173.0
162.90	375.0
164.00	118.0

Instrument :
MSVOA_W
ClientSampleId :
BFB37

165.00	230.0
171.80	209.0
174.00	40904.0
175.00	3252.0
176.00	40920.0
177.00	3747.0
178.00	343.0
179.00	559.0
180.10	128.0
189.00	105.0
191.00	2036.0
192.00	384.0
193.00	2469.0
194.00	297.0
195.00	244.0
203.00	211.0
205.00	484.0
206.10	148.0
207.00	1401.0
208.00	375.0
209.10	130.0
220.90	85.0
235.00	204.0
249.00	1556.0
250.00	509.0
251.00	677.0
252.00	149.0
252.90	69.0
265.00	1923.0
266.10	473.0
267.10	423.0
268.00	149.0
269.00	68.0
279.20	97.0
281.10	17584.0
282.00	6741.0
283.10	3989.0
284.10	796.0
285.10	312.0
286.00	52.0

Instrument :
MSVOA_W
ClientSampleId :
BFB37

m/z	Abundance
35.00	103.0
36.00	643.0
37.00	2877.0
38.00	2599.0
39.00	919.0
40.00	214.0
41.00	56.0
42.90	97.0
44.00	542.0
45.00	675.0
46.10	52.0
47.00	704.0
48.00	423.0
49.00	2663.0
50.00	11966.0
51.00	3838.0
51.90	241.0
53.00	54.0
55.00	118.0
56.00	1015.0
57.00	1778.0
58.10	136.0
59.10	212.0
59.90	691.0
61.00	3214.0
62.00	3181.0
63.10	2231.0
63.90	226.0
64.90	129.0
66.80	154.0
68.00	6881.0
69.00	7686.0
70.00	532.0
72.00	392.0
73.10	3803.0
74.00	11550.0
75.00	35064.0
76.00	3149.0
76.90	351.0
77.90	353.0
78.90	1877.0
79.90	444.0
81.00	2177.0
81.90	412.0
83.00	61.0
86.10	75.0
86.90	2590.0
87.90	2616.0
90.90	236.0
92.00	2138.0
93.00	3016.0
94.00	8725.0
95.00	72224.0
96.00	4656.0
97.10	124.0
103.00	218.0
103.90	345.0
104.90	174.0
105.80	332.0
107.00	107.0
110.10	153.0
110.90	234.0
111.80	86.0
112.90	86.0
114.90	139.0
115.90	325.0
117.00	575.0
118.00	479.0
118.90	574.0
125.00	380.0
125.50	197.0
126.10	168.0
127.90	339.0
128.80	142.0
129.90	319.0
130.90	211.0
133.00	1429.0
134.20	214.0
134.90	173.0
136.80	103.0
141.00	758.0
142.00	90.0
143.00	704.0
144.90	124.0
145.80	116.0
146.80	250.0
147.90	258.0
148.80	92.0
149.70	52.0
150.10	73.0
155.00	171.0
156.90	140.0

Instrument :
MSVOA_W
ClientSampleId :
BFB37

159.00	114.0
160.90	196.0
162.90	222.0
164.90	176.0
171.90	335.0
174.00	57992.0
175.00	4301.0
175.90	56512.0
177.00	4280.0
177.90	195.0
178.90	343.0
189.00	94.0
191.00	1413.0
192.00	252.0
193.00	1792.0
194.00	304.0
195.00	113.0
203.00	173.0
205.10	294.0
206.00	151.0
207.10	995.0
208.10	164.0
209.00	137.0
220.90	77.0
235.00	174.0
249.00	1107.0
250.00	356.0
251.00	547.0
252.00	162.0
252.90	92.0
265.10	1314.0
265.90	358.0
267.10	419.0
267.90	116.0
281.10	15693.0
282.10	4282.0
283.10	3263.0
284.00	628.0
285.00	125.0

Instrument :
MSVOA_W
ClientSampleId :
BFB37

m/z	Abundance
36.00	708.0
37.00	2512.0
38.10	2110.0
39.00	976.0
40.10	182.0
44.00	460.0
45.00	549.0
46.10	81.0
47.00	842.0
48.10	384.0
49.00	2374.0
50.10	10142.0
51.00	3360.0
52.00	181.0
55.10	192.0
56.10	998.0
57.10	1385.0
58.00	99.0
60.10	545.0
61.00	2717.0
62.00	2957.0
63.00	2051.0
64.10	200.0
65.00	169.0
67.00	152.0
68.00	6189.0
69.00	6326.0
70.10	467.0
71.90	342.0
73.00	3259.0
74.00	10874.0
75.00	31920.0
76.10	2663.0
76.90	475.0
77.70	214.0
79.00	1597.0
80.00	475.0
81.00	1779.0
82.00	360.0
86.90	2670.0
88.00	2740.0
90.90	179.0
92.00	1759.0
93.00	2950.0
94.00	7483.0
95.00	66888.0
96.10	4249.0
97.00	119.0
102.90	104.0
103.90	351.0
104.80	182.0
105.90	257.0
106.80	72.0
110.90	93.0
114.70	85.0
116.00	286.0
116.90	453.0
118.00	226.0
118.90	350.0
125.10	247.0
127.80	225.0
128.90	186.0
130.00	215.0
130.90	134.0
133.10	718.0
134.90	228.0
136.90	125.0
140.00	55.0
140.90	636.0
142.10	73.0
142.90	731.0
144.90	57.0
145.80	148.0
146.90	128.0
147.90	179.0
148.90	118.0
149.70	50.0
152.90	58.0
154.80	165.0
156.90	206.0
159.00	148.0
160.90	165.0
163.10	126.0
164.80	88.0
172.00	148.0
173.90	61064.0
175.00	4289.0
175.90	56112.0
177.00	3573.0
178.00	126.0
178.90	164.0
190.90	524.0

Instrument :
MSVOA_W
ClientSampleId :
BFB37

192.10	163.0
193.00	920.0
193.90	142.0
195.00	124.0
202.80	73.0
204.90	196.0
207.00	601.0
208.10	169.0
208.90	93.0
235.10	67.0
249.00	577.0
249.90	145.0
251.10	287.0
265.00	801.0
266.00	229.0
266.80	146.0
267.20	122.0
281.10	8818.0
282.00	2368.0
283.10	1328.0
284.10	404.0
285.20	112.0

Instrument :
MSVOA_W
ClientSampleId :
BFB37

m/z	Abundance
36.00	218.0
38.20	95.0
40.00	205.0
42.10	57.0
42.90	69.0
43.90	151.0
55.10	67.0
68.90	57.0
98.00	52.0

Instrument :
MSVOA_W
ClientSampleId :
BFB37