

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW020922\
 Data File : VW022078.D
 Acq On : 09 Feb 2022 10:24
 Operator : SY/VA
 Sample : N1364-11
 Misc : 5.06g/10mL/MSVOA_W/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 VHBLK001

Quant Time: Feb 10 04:51:23 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM020122SMA.M
 Quant Title : SFAM01.0
 QLast Update : Thu Feb 03 01:57:48 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.841	114	139051	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.633	117	130650	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.554	152	63952	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.349	65	39246	23.431	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	93.720%
7) Chloroethane-d5	2.885	69	30457	23.620	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	94.480%
11) 1,1-Dichloroethene-d2	4.019	63	50871	17.191	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	68.760%
21) 2-Butanone-d5	7.079	46	14170	47.697	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	95.400%
24) Chloroform-d	7.646	84	79373	24.535	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	98.120%
26) 1,2-Dichloroethane-d4	8.305	65	40915	25.420	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	101.680%
32) Benzene-d6	8.274	84	154366	23.381	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	93.520%
36) 1,2-Dichloropropane-d6	9.274	67	43318	24.121	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	96.480%
41) Toluene-d8	10.323	98	151286	23.120	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	92.480%
43) trans-1,3-Dichloroprop...	10.579	79	17068	21.701	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	86.800%
47) 2-Hexanone-d5	10.926	63	12738	48.787	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	97.580%
56) 1,1,2,2-Tetrachloroeth...	12.688	84	34567	23.600	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	94.400%
66) 1,2-Dichlorobenzene-d4	13.853	152	54621	23.676	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	94.720%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW020922\
Data File : VW022078.D
Acq On : 09 Feb 2022 10:24
Operator : SY/VA
Sample : N1364-11
Misc : 5.06g/10mL/MSVOA_W/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
VHBLK001

Quant Time: Feb 10 04:51:23 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM020122SMA.M
Quant Title : SFAM01.0
QLast Update : Thu Feb 03 01:57:48 2022
Response via : Initial Calibration

