

Data Path : Z:\VOASRV\HPCHEM1\MSVOA W\DATA\VW021219\
 Data File : VW008615.D
 Acq On : 12 Feb 2019 11:42
 Operator : SY/VA
 Sample : VW0212SBL01
 Misc : 5.00G/5ML/MSVOA W/SOIL
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 VW0212SBL01

Quant Time: Feb 13 00:33:20 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\82W012419S.M
 Quant Title : SW846 8260
 QLast Update : Fri Jan 25 11:07:41 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.95	168	410406	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.84	114	661138	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.63	117	602096	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.56	152	302823	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.30	65	220204	54.62	ug/l	0.00
Spiked Amount 50.000			Recovery	=	109.24%	
35) Dibromofluoromethane	7.88	113	206328	53.37	ug/l	0.00
Spiked Amount 50.000			Recovery	=	106.74%	
50) Toluene-d8	10.32	98	792687	51.13	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.26%	
62) 4-Bromofluorobenzene	12.62	95	295419	49.58	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.16%	
Target Compounds						
20) Methylene Chloride	4.90	84	11619	2.784	ug/l	Qvalue 88

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA W\DATA\VW021219\
Data File : VW008615.D
Acq On : 12 Feb 2019 11:42
Operator : SY/VA
Sample : VW0212SBL01
Misc : 5.00G/5ML/MSVOA W/SOIL
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
VW0212SBL01

Quant Time: Feb 13 00:33:20 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_W\METHOD\82W012419S.M
Quant Title : SW846 8260
QLast Update : Fri Jan 25 11:07:41 2019
Response via : Initial Calibration

