

Data Path : Z:\VOASRV\HPCHEM1\MSVOA W\DATA\VW021219\
 Data File : VW008619.D
 Acq On : 12 Feb 2019 13:38
 Operator : SY/VA
 Sample : K1454-01
 Misc : 6.39G/5ML/MSVOA W/SOIL
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 TP-1

Quant Time: Feb 13 00:41:02 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\82W012419S.M
 Quant Title : SW846 8260
 QLast Update : Fri Jan 25 11:07:41 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.94	168	397738	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.84	114	623944	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.63	117	525741	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.56	152	192427	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.30	65	218369	55.89	ug/l	0.00
Spiked Amount 50.000			Recovery	=	111.78%	
35) Dibromofluoromethane	7.88	113	201407	55.20	ug/l	0.00
Spiked Amount 50.000			Recovery	=	110.40%	
50) Toluene-d8	10.32	98	738714	50.49	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.98%	
62) 4-Bromofluorobenzene	12.62	95	222640	39.59	ug/l	0.00
Spiked Amount 50.000			Recovery	=	79.18%	
Target Compounds						
16) Acetone	4.12	43	33060	37.677	ug/l	99
20) Methylene Chloride	4.91	84	22580	5.583	ug/l	90
32) 1,1,1-Trichloroethane	7.87	97	28230	4.481	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA W\DATA\VW021219\
Data File : VW008619.D
Acq On : 12 Feb 2019 13:38
Operator : SY/VA
Sample : K1454-01
Misc : 6.39G/5ML/MSVOA W/SOIL
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
TP-1

Quant Time: Feb 13 00:41:02 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_W\METHOD\82W012419S.M
Quant Title : SW846 8260
QLast Update : Fri Jan 25 11:07:41 2019
Response via : Initial Calibration

