

Data Path : Z:\VOASRV\HPCHEM1\MSVOA W\DATA\VW021219\
 Data File : VW008633.D
 Acq On : 12 Feb 2019 19:52
 Operator : SY/VA
 Sample : K1343-19
 Misc : 6.52G/5ML/MSVOA W/SOIL
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 JC-02-021119-J

Quant Time: Feb 13 01:32:46 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\82W012419S.M
 Quant Title : SW846 8260
 QLast Update : Fri Jan 25 11:07:41 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.95	168	406285	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.84	114	673172	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.63	117	612643	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.56	152	282734	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.30	65	233337	58.46	ug/l	0.00
Spiked Amount 50.000			Recovery	=	116.92%	
35) Dibromofluoromethane	7.88	113	209614	53.25	ug/l	0.00
Spiked Amount 50.000			Recovery	=	106.50%	
50) Toluene-d8	10.32	98	805069	51.00	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.00%	
62) 4-Bromofluorobenzene	12.62	95	295514	48.71	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.42%	
Target Compounds						
16) Acetone	4.12	43	68949	76.924	ug/l	96
20) Methylene Chloride	4.92	84	15723	3.806	ug/l #	88
52) Toluene	10.38	92	20828	1.788	ug/l	97
68) m/p-Xylenes	11.83	106	12401	1.390	ug/l	99
84) 1,2,4-Trimethylbenzene	13.25	105	60441	3.375	ug/l	99
95) Naphthalene	15.37	128	41295	4.006	ug/l #	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA W\DATA\VW021219\
Data File : VW008633.D
Acq On : 12 Feb 2019 19:52
Operator : SY/VA
Sample : K1343-19
Misc : 6.52G/5ML/MSVOA W/SOIL
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
JC-02-021119-J

Quant Time: Feb 13 01:32:46 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_W\METHOD\82W012419S.M
Quant Title : SW846 8260
QLast Update : Fri Jan 25 11:07:41 2019
Response via : Initial Calibration

