

Quantitation Report (Qedit)




Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW021221\
 Data File : VW018103.D
 Acq On : 12 Feb 2021 12:23
 Operator : SY/VA
 Sample : M1371-13
 Misc : 7.01G/10ML/MSVOA_W/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 DBH24

Manual Integrations
 APPROVED

MMDadoda
 2/16/2021 12:57:03 PM

Quant Time: Feb 12 22:44:02 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SOM2WLM012921S.M
 Quant Title : VOC Analysis
 QLast Update : Fri Feb 12 22:41:34 2021
 Response via : Initial Calibration



TIC: VW018103.D\data.ms

(13) Acetone (T)

4.123min (+ 0.006) 0.58 ug/L

response 463

Ion	Exp%	Act%
43.00	100.00	100.00
58.00	34.20	39.31
0.00	0.00	0.00
0.00	0.00	0.00

Quantitation Report (Qedit)




Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW021221\
 Data File : VW018103.D
 Acq On : 12 Feb 2021 12:23
 Operator : SY/VA
 Sample : M1371-13
 Misc : 7.01G/10ML/MSVOA_W/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 DBH24

Manual Integrations
 APPROVED

MMDadoda
 2/16/2021 12:57:03 PM

Quant Time: Feb 12 22:44:02 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SOM2WLM012921S.M
 Quant Title : VOC Analysis
 QLast Update : Fri Feb 12 22:41:34 2021
 Response via : Initial Calibration



TIC: VW018103.D\data.ms

(13) Acetone (T)

4.123min (+ 0.006) 1.09 ug/L m

> 02/17/21 sy

response 874

Ion	Exp%	Act%
43.00	100.00	100.00
58.00	34.20	20.82
0.00	0.00	0.00
0.00	0.00	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW021221\
 Data File : VW018103.D
 Acq On : 12 Feb 2021 12:23
 Operator : SY/VA
 Sample : M1371-13
 Misc : 7.01G/10ML/MSVOA_W/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 DBH24

Manual Integrations
 APPROVED

MMDadoda
 2/16/2021 12:57:03 PM

Quant Time: Feb 12 22:44:02 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SOM2WLM012921S.M
 Quant Title : VOC Analysis
 QLast Update : Fri Feb 12 22:41:34 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.841	114	325740	25.00	ug/L	0.00
28) Chlorobenzene-d5	11.634	117	229366	25.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	13.560	152	68494	25.00	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.355	65	63126	15.59	ug/L	0.00
Spiked Amount 25.000	Range 30 - 150		Recovery =	62.36%		
7) Chloroethane-d5	2.891	69	56946	17.72	ug/L	0.00
Spiked Amount 25.000	Range 30 - 150		Recovery =	70.88%		
10) 1,1-Dichloroethene-d2	4.019	63	102641	12.68	ug/L	0.00
Spiked Amount 25.000	Range 45 - 110		Recovery =	50.72%		
20) 2-Butanone-d5	7.074	46	31745	29.38	ug/L	0.00
Spiked Amount 50.000	Range 20 - 135		Recovery =	58.76%		
24) Chloroform-d	7.647	84	160349	18.76	ug/L	0.00
Spiked Amount 25.000	Range 40 - 150		Recovery =	75.04%		
26) 1,2-Dichloroethane-d4	8.305	65	83411	18.32	ug/L	0.00
Spiked Amount 25.000	Range 70 - 130		Recovery =	73.28%		
29) Benzene-d6	8.275	84	306767	24.36	ug/L	0.00
Spiked Amount 25.000	Range 20 - 135		Recovery =	97.44%		
33) 1,2-Dichloropropane-d6	9.274	67	88593	25.15	ug/L	0.00
Spiked Amount 25.000	Range 70 - 120		Recovery =	100.60%		
37) Toluene-d8	10.323	98	249469	20.82	ug/L	0.00
Spiked Amount 25.000	Range 30 - 130		Recovery =	83.28%		
38) trans-1,3-Dichloroprop...	10.579	79	32506	20.36	ug/L	0.00
Spiked Amount 25.000	Range 30 - 135		Recovery =	81.44%		
39) 2-Hexanone-d5	10.927	63	25775	38.27	ug/L	0.00
Spiked Amount 50.000	Range 20 - 135		Recovery =	76.54%		
48) 1,1,2,2-Tetrachloroeth...	12.694	84	59691	19.29	ug/L	0.00
Spiked Amount 25.000	Range 45 - 120		Recovery =	77.16%		
61) 1,2-Dichlorobenzene-d4	13.853	152	48793	19.97	ug/L	0.00
Spiked Amount 25.000	Range 75 - 120		Recovery =	79.88%		
Target Compounds						
13) Acetone	4.123	43	874m	1.09	ug/L	Qvalue 91
16) Methylene chloride	4.915	84	6130	1.17	ug/L	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW021221\
Data File : VW018103.D
Acq On : 12 Feb 2021 12:23
Operator : SY/VA
Sample : M1371-13
Misc : 7.01G/10ML/MSVOA_W/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampled :
DBH24

Manual Integrations
APPROVED

MMDadoda
2/16/2021 12:57:03 PM

Quant Time: Feb 12 22:44:02 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SOM2WLM012921S.M
Quant Title : VOC Analysis
QLast Update : Fri Feb 12 22:41:34 2021
Response via : Initial Calibration

