

Data Path : Z:\VOASRV\HPCHEM1\MSVOA W\DATA\VW021319\
 Data File : VW008648.D
 Acq On : 13 Feb 2019 20:37
 Operator : SY/VA
 Sample : K1343-27
 Misc : 6.74G/5ML/MSVOA W/SOIL
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 JC-02-021119-N

Quant Time: Feb 14 05:33:39 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\82W021319S.M
 Quant Title : SW846 8260
 QLast Update : Thu Feb 14 04:06:50 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.95	168	433312	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.84	114	644604	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.63	117	564716	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.56	152	272578	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.30	65	230326	51.13	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.26%	
35) Dibromofluoromethane	7.88	113	203011	50.72	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.44%	
50) Toluene-d8	10.32	98	770773	49.95	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.90%	
62) 4-Bromofluorobenzene	12.62	95	265091	45.61	ug/l	0.00
Spiked Amount 50.000			Recovery	=	91.22%	
Target Compounds						
16) Acetone	4.12	43	47018	45.866	ug/l	100
17) Carbon Disulfide	4.37	76	15694	1.229	ug/l #	94
20) Methylene Chloride	4.91	84	43772	2.872	ug/l	91
80) 1,3,5-Trimethylbenzene	12.94	105	21129	1.253	ug/l	100
84) 1,2,4-Trimethylbenzene	13.25	105	61720	3.561	ug/l	100
89) n-Butylbenzene	13.83	91	19456	1.073	ug/l #	77
95) Naphthalene	15.37	128	24440	2.286	ug/l #	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA W\DATA\VW021319\
Data File : VW008648.D
Acq On : 13 Feb 2019 20:37
Operator : SY/VA
Sample : K1343-27
Misc : 6.74G/5ML/MSVOA W/SOIL
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
JC-02-021119-N

Quant Time: Feb 14 05:33:39 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_W\METHOD\82W021319S.M
Quant Title : SW846 8260
QLast Update : Thu Feb 14 04:06:50 2019
Response via : Initial Calibration

