




Data File : VW014982.D
Acq On : 13 Feb 2020 16:03
Operator : SY/VA
Sample : MDL01
Misc : 5.00G/10ML/MSVOA_W/SOIL
ALS Vial : 1 Sample Multiplier: 1

Quant Time: Feb 14 04:59:21 2020
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\SFAMWLM021320SMA.M
Quant Title : SFAM01.0
QLast Update : Fri Feb 14 04:50:37 2020
Response via : Initial Calibration



TIC: VW014982.D

(5) Vinyl chloride (T)

2.477min (+0.110) 0.01ug/L

response 175

Ion	Exp%	Act%
62.00	100	100
64.00	33.70	42.35
0.00	0.00	0.00
0.00	0.00	0.00