




Data File : VW014982.D
Acq On : 13 Feb 2020 16:03
Operator : SY/VA
Sample : MDL01
Misc : 5.00G/10ML/MSVOA_W/SOIL
ALS Vial : 1 Sample Multiplier: 1

Quant Time: Feb 14 04:59:21 2020
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\SFAMWLM021320SMA.M
Quant Title : SFAM01.0
QLast Update : Fri Feb 14 04:50:37 2020
Response via : Initial Calibration



TIC: VW014982.D

(10) 1,1,2-Trichloro-1,2,2-trifluoroethane (T)

4.050min (-0.018) 1.03ug/L

response 11759

Ion	Exp%	Act%
101.00	100	100
85.00	45.60	79.37#
151.00	62.80	119.47#
0.00	0.00	0.00