




Data File : VW014982.D
Acq On : 13 Feb 2020 16:03
Operator : SY/VA
Sample : MDL01
Misc : 5.00G/10ML/MSVOA_W/SOIL
ALS Vial : 1 Sample Multiplier: 1

Quant Time: Feb 14 04:59:21 2020
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\SFAMWLM021320SMA.M
Quant Title : SFAM01.0
QLast Update : Fri Feb 14 04:50:37 2020
Response via : Initial Calibration



TIC: VW014982.D

(44) trans-1,3-Dichloropropene (T)

10.604min (0.000) 2.32ug/L

response 36441

Ion	Exp%	Act%
75.00	100	100
77.00	32.40	33.26
0.00	0.00	0.00
0.00	0.00	0.00