

Data Path : Z:\VOASRV\HPCHEM1\MSVOA W\DATA\VW021519\
 Data File : VW008717.D
 Acq On : 16 Feb 2019 07:31
 Operator : SY/VA
 Sample : K1513-01
 Misc : 5.00G/5ML/MSVOA W/SOIL
 ALS Vial : 49 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 FLOORS

Quant Time: Feb 18 08:09:42 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\82W021319S.M
 Quant Title : SW846 8260
 QLast Update : Thu Feb 14 04:06:50 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.95	168	277366	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.84	114	426589	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.63	117	381398	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.56	152	187217	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.31	65	153714	53.31	ug/l	0.00
Spiked Amount 50.000			Recovery	=	106.62%	
35) Dibromofluoromethane	7.88	113	43271	16.34	ug/l	0.00
Spiked Amount 50.000			Recovery	=	32.68%	
50) Toluene-d8	10.32	98	503838	49.34	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.68%	
62) 4-Bromofluorobenzene	12.62	95	178073	46.30	ug/l	0.00
Spiked Amount 50.000			Recovery	=	92.60%	
Target Compounds						
16) Acetone	4.12	43	52387	79.836	ug/l	89
25) 2-Butanone	7.17	43	15840	17.095	ug/l	90
59) 2-Hexanone	10.97	43	9153	6.153	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA W\DATA\VW021519\
Data File : VW008717.D
Acq On : 16 Feb 2019 07:31
Operator : SY/VA
Sample : K1513-01
Misc : 5.00G/5ML/MSVOA W/SOIL
ALS Vial : 49 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
FLOORS

Quant Time: Feb 18 08:09:42 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_W\METHOD\82W021319S.M
Quant Title : SW846 8260
QLast Update : Thu Feb 14 04:06:50 2019
Response via : Initial Calibration

