

Data Path : Z:\VOASRV\HPCHEM1\MSVOA W\DATA\VW021519\
 Data File : VW008718.D
 Acq On : 16 Feb 2019 07:58
 Operator : SY/VA
 Sample : K1513-03
 Misc : 5.10G/5ML/MSVOA W/SOIL
 ALS Vial : 50 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 WALL-ONE

Quant Time: Feb 18 08:11:09 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\82W021319S.M
 Quant Title : SW846 8260
 QLast Update : Thu Feb 14 04:06:50 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.95	168	347950	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.84	114	527740	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.63	117	458198	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.56	152	228887	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.30	65	176844	48.89	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.78%	
35) Dibromofluoromethane	7.88	113	164128	50.09	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.18%	
50) Toluene-d8	10.32	98	617703	48.90	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.80%	
62) 4-Bromofluorobenzene	12.62	95	218297	45.88	ug/l	0.00
Spiked Amount 50.000			Recovery	=	91.76%	
Target Compounds						
16) Acetone	4.12	43	7364	8.946	ug/l #	84
20) Methylene Chloride	4.91	84	160501	43.282	ug/l	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA W\DATA\VW021519\
Data File : VW008718.D
Acq On : 16 Feb 2019 07:58
Operator : SY/VA
Sample : K1513-03
Misc : 5.10G/5ML/MSVOA W/SOIL
ALS Vial : 50 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
WALL-ONE

Quant Time: Feb 18 08:11:09 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_W\METHOD\82W021319S.M
Quant Title : SW846 8260
QLast Update : Thu Feb 14 04:06:50 2019
Response via : Initial Calibration

