

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW021524\
 Data File : VW028021.D
 Acq On : 15 Feb 2024 15:43
 Operator : SY/MD
 Sample : P1469-12
 Misc : 5.87g/10mL/MSVOA_W/SOIL/A
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 BH9X3

Quant Time: Feb 15 20:54:21 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM020524SMA.M
 Quant Title : SFAM01.0
 QLast Update : Thu Feb 15 20:52:49 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.843	114	109789	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.629	117	105962	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.556	152	51778	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.357	65	21797	16.951	ug/L	-0.01
Spiked Amount	25.000	Range	30 - 150	Recovery	=	67.800%
7) Chloroethane-d5	2.893	69	21542	22.087	ug/L	-0.01
Spiked Amount	25.000	Range	30 - 150	Recovery	=	88.360%
11) 1,1-Dichloroethene-d2	4.021	65	10033	18.235	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	72.920%
21) 2-Butanone-d5	7.081	46	13163	65.881	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	131.760%
24) Chloroform-d	7.648	84	59540	23.193	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	92.760%
26) 1,2-Dichloroethane-d4	8.307	65	31889	27.504	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	110.000%
32) Benzene-d6	8.276	84	116725	20.589	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	82.360%
36) 1,2-Dichloropropane-d6	9.270	67	34228	22.267	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	89.080%
41) Toluene-d8	10.319	98	105797	19.504	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	78.000%
43) trans-1,3-Dichloroprop...	10.581	79	14064	22.248	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	89.000%
47) 2-Hexanone-d5	10.922	63	11950	57.154	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	114.300%
56) 1,1,2,2-Tetrachloroeth...	12.690	84	32940	28.671	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	114.680%
66) 1,2-Dichlorobenzene-d4	13.848	152	40751	23.155	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	92.640%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW021524\
Data File : VW028021.D
Acq On : 15 Feb 2024 15:43
Operator : SY/MD
Sample : P1469-12
Misc : 5.87g/10mL/MSVOA_W/SOIL/A
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
BH9X3

Quant Time: Feb 15 20:54:21 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM020524SMA.M
Quant Title : SFAM01.0
QLast Update : Thu Feb 15 20:52:49 2024
Response via : Initial Calibration

