


Data Path : Z:\VOASRV\HPCHEM1\MSVOA W\DATA\VW021819\
Data File : VW008720.D
Acq On : 18 Feb 2019 09:37
Operator : SY/VA
Sample : BFB
Misc : 5.00G/5ML/MSVOA W/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\82W021319S.M
Title : SW846 8260
Last Update : Thu Feb 14 04:06:50 2019



AutoFind: Scans 1757, 1758, 1759; Background Corrected with Scan 1745

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	20.5	12939	PASS
75	95	30	60	52.1	32781	PASS
95	95	100	100	100.0	62978	PASS
96	95	5	9	6.2	3895	PASS
173	174	0.00	2	1.5	821	PASS
174	95	50	100	87.0	54808	PASS
175	174	5	9	5.3	2912	PASS
176	174	95	101	98.0	53725	PASS
177	176	5	9	7.0	3752	PASS