

Data Path : Z:\VOASRV\HPCHEM1\MSVOA W\DATA\VW021919\
 Data File : VW008749.D
 Acq On : 19 Feb 2019 17:26
 Operator : SY/VA
 Sample : K1349-09
 Misc : 6.25G/5ML/MSVOA W/SOIL
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 OR-02-021919-A

Quant Time: Feb 20 07:08:42 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\82W021319S.M
 Quant Title : SW846 8260
 QLast Update : Thu Feb 14 04:06:50 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.95	168	367932	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.84	114	559284	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.63	117	470326	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.56	152	201795	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.30	65	193153	50.50	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.00%	
35) Dibromofluoromethane	7.88	113	179341	51.64	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.28%	
50) Toluene-d8	10.32	98	656430	49.03	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.06%	
62) 4-Bromofluorobenzene	12.62	95	208852	41.42	ug/l	0.00
Spiked Amount 50.000			Recovery	=	82.84%	
Target Compounds						
16) Acetone	4.13	43	35314	40.570	ug/l	98
20) Methylene Chloride	4.90	84	37011	2.825	ug/l	91
52) Toluene	10.38	92	128843	13.014	ug/l	98
68) m/p-Xylenes	11.83	106	17024	2.423	ug/l	96
69) o-Xylene	12.16	106	8854	1.333	ug/l	97
80) 1,3,5-Trimethylbenzene	12.94	105	16070	1.287	ug/l	98
84) 1,2,4-Trimethylbenzene	13.25	105	44399	3.460	ug/l	98
95) Naphthalene	15.37	128	81153	10.251	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA W\DATA\VW021919\
Data File : VW008749.D
Acq On : 19 Feb 2019 17:26
Operator : SY/VA
Sample : K1349-09
Misc : 6.25G/5ML/MSVOA W/SOIL
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
OR-02-021919-A

Quant Time: Feb 20 07:08:42 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_W\METHOD\82W021319S.M
Quant Title : SW846 8260
QLast Update : Thu Feb 14 04:06:50 2019
Response via : Initial Calibration

